



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB88054

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/13/2017	16:05
2	CAL2	1	Water			7.01	06/13/2017	16:07
3	CAL3	1	Water			10.02	06/13/2017	16:09
4	ICV	1	Water			7.01	06/13/2017	16:11
5	CCV1	1	Water			2.02	06/13/2017	16:13
6	I3643-03	1	Solid	20.02	20	8.13	06/13/2017	16:15
7	I3643-03DUP	1	Solid	20.04	20	8.14	06/13/2017	16:17
8	I3660-03	1	Solid	20.02	20	8.05	06/13/2017	16:19
9	I3660-07	1	Solid	20.00	20	7.54	06/13/2017	16:20
10	I3660-11	1	Solid	20.03	20	6.72	06/13/2017	16:22
11	I3660-15	1	Solid	20.04	20	5.45	06/13/2017	16:25
12	I3660-19	1	Solid	20.00	20	6.91	06/13/2017	16:26
13	I3665-05	1	Solid	20.03	20	8.46	06/13/2017	16:27
14	I3665-10	1	Solid	20.02	20	9.04	06/13/2017	16:30
15	I3670-04	1	Solid	20.01	20	5.13	06/13/2017	16:32
16	CCV2	1	Water			12.01	06/13/2017	16:33

LB 88054

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	CORROSIONITY I3643			WorkList ID :	99522	Date :	6/13/2017 3:24:21 PM		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2017	Solid	I3643-03	Corrosivity	Cool 4 deg C	PSEG01	L51	RO-VNJ217	06/12/2017	9045C
06/16/2017	Solid	I3660-03	Corrosivity	Cool 4 deg C	PSEG01	M51	RO-132	06/13/2017	9045C
06/16/2017	Solid	I3660-07	Corrosivity	Cool 4 deg C	PSEG01	M51	RO-153	06/13/2017	9045C
06/16/2017	Solid	I3660-11	Corrosivity	Cool 4 deg C	PSEG01	M51	RO-351	06/13/2017	9045C
06/16/2017	Solid	I3660-15	Corrosivity	Cool 4 deg C	PSEG01	M51	RO-350	06/13/2017	9045C
06/16/2017	Solid	I3660-19	Corrosivity	Cool 4 deg C	PSEG01	M51	RO-72-11941	06/13/2017	9045C
	Solid	I3665-05	Corrosivity	Cool 4 deg C	DEWB01	M42	HA3-WC1	06/13/2017	9045C
	Solid	I3665-10	Corrosivity	Cool 4 deg C	DEWB01	M42	HA3-WC2	06/13/2017	9045C
06/20/2017	Solid	I3670-04	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-CLINTON-AVE-1-2	06/13/2017	9045C

Date/Time 06-13-17 7:00 PM
Received by: J.C.
Relinquished by: J.V.

Date/Time 06-13-17 3:24 PM
Received by: J.V.
Relinquished by: J.C.