


7470A

LB88059
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37206



A= 0.0000e+000

B= 2.7173e-004

C= -3.0144e-003

Rho= 0.9995701

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	54	0.000	54				
0.2	0.200	0.227	0.027	845	0.0 %	845				
2.5	2.500	2.457	-0.043	9054	0.0 %	9054				
5.0	5.000	5.108	0.108	18809	0.0 %	18809				
7.5	7.500	7.289	-0.211	26836	0.0 %	26836				
10.0	10.000	10.107	0.107	37206	0.0 %	37206				

LB88059 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Jun 2017 10:55:33

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	54	-	0.00007470A	PPB	14 Jun 2017 11:00:48	S	Std
0.2 - 1	50.2	845	-	0.20007470A	PPB	14 Jun 2017 11:02:57	S	Std
2.5 - 1	52.5	9054	-	2.50007470A	PPB	14 Jun 2017 11:05:05	S	Std
5.0 - 1	55	18809	-	5.00007470A	PPB	14 Jun 2017 11:07:13	S	Std
7.5 - 1	57.5	26836	-	7.50007470A	PPB	14 Jun 2017 11:09:22	S	Std
10.0 - 1	510	37206	-	10.00007470A	PPB	14 Jun 2017 11:11:28	S	Std
ICV97 - 1	ICV97	14571	3.9564	-7470A	PPB	14 Jun 2017 11:14:03	U	SMPL
ICB97 - 1	ICB97	46	0.0095	-7470A	PPB	14 Jun 2017 11:16:10	U	SMPL
CCV40 - 1	CCV40	19020	5.1654	-7470A	PPB	14 Jun 2017 11:18:16	U	SMPL
CCB40 - 1	CCB40	19	0.0021	-7470A	PPB	14 Jun 2017 11:20:22	U	SMPL
CRA - 1	CRA	832	0.2231	-7470A	PPB	14 Jun 2017 11:22:29	U	SMPL
HighStd - 1	HighStd	36140	9.8175	-7470A	PPB	14 Jun 2017 11:24:36	U	SMPL
ChkStd - 1	ChkStd	27218	7.3931	-7470A	PPB	14 Jun 2017 11:26:42	U	SMPL
PB99741BL - 1	PBW	-46	-0.0155	-7470A	PPB	14 Jun 2017 11:28:50	U	SMPL
PB99741BS - 1	LCSW	14446	3.9225	-7470A	PPB	14 Jun 2017 11:30:58	U	SMPL
PB99741TB - 1	PB99701TB	-24	-0.0095	-7470A	PPB	14 Jun 2017 11:33:05	U	SMPL
I3628-03 - 1	COMPOSITE-1	28	0.0046	-7470A	PPB	14 Jun 2017 11:35:14	U	SMPL
I3628-03DUP - 1	COMPOSITE-1DUP	25	0.0038	-7470A	PPB	14 Jun 2017 11:37:21	U	SMPL
I3628-03MS - 1	COMPOSITE-1MS	14089	3.8255	-7470A	PPB	14 Jun 2017 11:39:30	U	SMPL
I3628-03MSD - 1	COMPOSITE-1MSD	14211	3.8586	-7470A	PPB	14 Jun 2017 11:41:37	U	SMPL
CCV41 - 1	CCV41	19168	5.2056	-7470A	PPB	14 Jun 2017 11:43:44	U	SMPL
CCB41 - 1	CCB41	37	0.0070	-7470A	PPB	14 Jun 2017 11:45:51	U	SMPL
I3629-03 - 1	CLINTON-AVE-SP-COMP-1	47	0.0098	-7470A	PPB	14 Jun 2017 11:47:58	U	SMPL
I3629-07 - 1	CLINTON-AVE-SP-COMP-2	48	0.0100	-7470A	PPB	14 Jun 2017 11:50:05	U	SMPL
I3630-07 - 1	TP-7	22	0.0030	-7470A	PPB	14 Jun 2017 11:52:12	U	SMPL
I3630-08 - 1	TP-8	34	0.0062	-7470A	PPB	14 Jun 2017 11:54:19	U	SMPL
I3630-09 - 1	TP-9	33	0.0060	-7470A	PPB	14 Jun 2017 11:56:26	U	SMPL
I3640-05 - 1	NF-DRUMS-STONE-COMP	96	0.0231	-7470A	PPB	14 Jun 2017 11:58:34	U	SMPL
I3640-10 - 1	NF-DRUMS-DEBRIS-COMP	93	0.0223	-7470A	PPB	14 Jun 2017 12:00:42	U	SMPL
I3628-03LX5 - 1		16	0.0013	-7470A	PPB	14 Jun 2017 12:02:50	U	SMPL
I3628-03A - 1		14975	4.0662	-7470A	PPB	14 Jun 2017 12:04:59	U	SMPL
CCV42 - 1	CCV42	18776	5.0991	-7470A	PPB	14 Jun 2017 12:07:05	U	SMPL
CCB42 - 1	CCB42	25	0.0038	-7470A	PPB	14 Jun 2017 12:09:12	U	SMPL