



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88069

Temp.: 20.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/13/2017	16:20
2	CAL2	1	Water			7.00	06/13/2017	16:21
3	CAL3	1	Water			10.03	06/13/2017	16:22
4	ICV	1	Water			7.01	06/13/2017	16:25
5	CCV1	1	Water			2.01	06/13/2017	16:29
6	I3644-01	1	Water			8.38	06/13/2017	16:30
7	I3644-01DUP	1	Water			8.39	06/13/2017	16:31
8	CCV2	1	Water			12.01	06/13/2017	16:33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH13644

WorkList ID : 99523

Date : 6/13/2017 4:14:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/13/2017	Water	I3644-01✓	pH	Cool 4 deg C	PSEG01	K42	DRUMS-M47409-COMP	06/13/2017	9045C

Date/Time 06/13/17 4:14 PM
 Received by: JP
 Relinquished by: JP

Date/Time 06/13/17 5:30 PM
 Received by: JP
 Relinquished by: JP