


ISM02.4

LB88089
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43374



A= 0.0000e+000

B= 2.3201e-004

C= -8.5044e-003

Rho= 0.9999533

Accept=Accepted

*slope
= 1/2148 cent*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	0.006	0.006	61	0.000	61					1
0.2	0.200	0.219	0.019	980	0.0 %	980					9
2.5	2.500	2.509	0.009	10851	0.0 %	10851					0
5.0	5.000	4.968	-0.032	21448	0.0 %	21448					1
7.5	7.500	7.444	-0.056	32122	0.0 %	32122					1
10.0	10.000	10.055	0.055	43374	0.0 %	43374					1

AB
6/14/17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	61	-		0 ISM02.4	PPB	6/14/2017 16:24	Std
	0.2 S0.2	980	-		0.2 ISM02.4	PPB	6/14/2017 16:26	Std
	2.5 S2.5	10851	-		2.5 ISM02.4	PPB	6/14/2017 16:30	Std
	5 S5	21448	-		5 ISM02.4	PPB	6/14/2017 16:33	Std
	7.5 S7.5	32122	-		7.5 ISM02.4	PPB	6/14/2017 16:35	Std
	10 S10	43374	-		10 ISM02.4	PPB	6/14/2017 16:37	Std
ICV	ICV	16613	3.8459 -		ISM02.4	PPB	6/14/2017 16:40	SMPL
ICB	ICB	-4	-0.0094 -		ISM02.4	PPB	6/14/2017 16:42	SMPL
CCV047	CCV047	21573	4.9966 -		ISM02.4	PPB	6/14/2017 16:44	SMPL
CCB047	CCB047	1	-0.0083 -		ISM02.4	PPB	6/14/2017 16:47	SMPL
PB99777BL	PBW003	12	-0.0057 -		ISM02.4	PPB	6/14/2017 16:49	SMPL
I3494-01	MBDC78	150	0.0263 -		ISM02.4	PPB	6/14/2017 16:51	SMPL
I3494-02	MBDC52	-63	-0.0231 -		ISM02.4	PPB	6/14/2017 16:53	SMPL
I3494-03	MBDC52D	-65	-0.0236 -		ISM02.4	PPB	6/14/2017 16:56	SMPL
I3494-04	MBDC52S	4337	0.9977 -		ISM02.4	PPB	6/14/2017 16:58	SMPL
I3494-05	MBDC68	-85	-0.0282 -		ISM02.4	PPB	6/14/2017 17:00	SMPL
I3494-06	MBDC82	-63	-0.0231 -		ISM02.4	PPB	6/14/2017 17:02	SMPL
I3494-07	MBDC72	-41	-0.018 -		ISM02.4	PPB	6/14/2017 17:05	SMPL
I3494-08	MBDCA6	-80	-0.0271 -		ISM02.4	PPB	6/14/2017 17:07	SMPL
I3494-09	MBDC58	-56	-0.0215 -		ISM02.4	PPB	6/14/2017 17:09	SMPL
I3494-10	MBDC85	-54	-0.021 -		ISM02.4	PPB	6/14/2017 17:12	SMPL
I3494-11	MBDC61	-58	-0.022 -		ISM02.4	PPB	6/14/2017 17:14	SMPL
I3494-12	MBDC59	-64	-0.0234 -		ISM02.4	PPB	6/14/2017 17:16	SMPL
I3494-13	MBDC60	-35	-0.0166 -		ISM02.4	PPB	6/14/2017 17:18	SMPL
I3494-14	MBDC62	-24	-0.0141 -		ISM02.4	PPB	6/14/2017 17:21	SMPL
CCV048	CCV048	21431	4.9637 -		ISM02.4	PPB	6/14/2017 17:23	SMPL
CCB048	CCB048	13	-0.0055 -		ISM02.4	PPB	6/14/2017 17:25	SMPL