



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88093

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/14/2017	17:40
2	CAL2	1	Water			7.01	06/14/2017	17:42
3	CAL3	1	Water			10.03	06/14/2017	17:43
4	ICV	1	Water			7.01	06/14/2017	17:45
5	CCV1	1	Water			2.01	06/14/2017	17:46
6	I3680-03	1	Solid	20.02	20	8.81	06/14/2017	17:47
7	I3680-03DUP	1	Solid	20.03	20	8.82	06/14/2017	17:48
8	I3681-06	1	Solid	20.04	20	8.58	06/14/2017	17:49
9	I3684-03	1	Solid	20.02	20	9.32	06/14/2017	17:51
10	I3684-07	1	Solid	20.03	20	8.88	06/14/2017	17:52
11	I3689-07	1	Solid	20.01	20	7.20	06/14/2017	17:53
12	I3689-14	1	Solid	20.03	20	6.55	06/14/2017	17:55
13	CCV2	1	Water			12.02	06/14/2017	18:00

21388093

WORKLIST(Hardcopy Internal Chain)

WorkList Name :		CORROSIONITY I3680		WorkList ID :		99580		Date :		6/14/2017 2:19:07 PM	
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method		
06/16/2017	Solid	I3680-03	Corrosivity	Cool 4 deg C	PSEG05	M31	TR-05-061317	06/13/2017	9045C		
06/16/2017	Solid	I3681-06	Corrosivity	Cool 4 deg C	PSEG01	M41	RO-272031-72-11978-COMP	06/13/2017	9045C		
06/16/2017	Solid	I3684-03	Corrosivity	Cool 4 deg C	PSEG01	M63	ROCKY-ROAD-COMP	06/13/2017	9045C		
06/16/2017	Solid	I3684-07	Corrosivity	Cool 4 deg C	PSEG01	M63	SOIL-ROAD-COMP	06/13/2017	9045C		
06/19/2017	Solid	I3689-07	Corrosivity	Cool 4 deg C	PSEG01	N12	BAYWAY-SWITCH-A	06/14/2017	9045C		
06/19/2017	Solid	I3689-14	Corrosivity	Cool 4 deg C	PSEG01	N12	BAYWAY-SWITCH-B	06/14/2017	9045C		

Date/Time 06/14/17 3:32 PM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 06/14/17 7:00 AM
Received by: [Signature]
Relinquished by: [Signature]