



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88094

Temp.: 20.2 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/14/2017	17:10
2	CAL2	1	Water			7.01	06/14/2017	17:11
3	CAL3	1	Water			10.03	06/14/2017	17:13
4	ICV	1	Water			7.00	06/14/2017	17:15
5	CCV1	1	Water			2.02	06/14/2017	17:16
6	I3661-04	1	Water			7.25	06/14/2017	17:19
7	I3691-02	1	Water			7.36	06/14/2017	17:20
8	I3691-02DUP	1	Water			7.37	06/14/2017	17:22
9	CCV2	1	Water			12.01	06/14/2017	17:25

WORKLIST(Hardcopy Internal Chain)

LB 88094

WorkList Name : CORROSIONITY 13661

WorkList ID : 99581

Date : 6/14/2017 2:20:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/14/2017	Water	13661-04	Corrosivity	Cool 4 deg C	PSEG01	M62	FT-265434-265578-COMP	06/13/2017	9040C
06/14/2017	Water	13691-02	Corrosivity	Cool 4 deg C	PSEG01	M21	A1013	06/14/2017	9040C

Date/Time 06-14-17 2:20:00
 Received by: JH
 Relinquished by: CL

Date/Time 06-14-17 2:20:00
 Received by: CL
 Relinquished by: JH