



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB88096

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, 5N sulfuric acid, Hexavalent Chromium ICV Std., etc.

Intercept: 0.0001 Slope: 0.7807 Regression: 0.999955

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Indiana

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB88096

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		2.31	0.000	0.390	0.390	0.499	06/14/2017	13:03
2	ICB		1	100	100		2.48	0.000	0.000	0.000	0.000	06/14/2017	13:04
3	CCV1	0.5	1	100	100		1.90	0.000	0.392	0.392	0.502	06/14/2017	18:00
4	CCB1		1	100	100		2.07	0.000	0.000	0.000	0.000	06/14/2017	18:00
5	PB99793BL		1	2.55	100	7.14	1.76	0.000	0.000	0.000	0.000	06/14/2017	18:01
6	PB99793BS	20	1	2.50	100	7.23	2.09	0.000	0.388	0.388	0.497	06/14/2017	18:01
7	I3566-02		1	2.51	100	7.09	2.44	0.012	0.012	0.000	0.000	06/14/2017	18:02
8	I3566-04		1	2.52	100	7.55	2.39	0.016	0.016	0.000	0.000	06/14/2017	18:02
9	I3566-06		1	2.50	100	7.19	2.21	0.019	0.019	0.000	0.000	06/14/2017	18:03
10	I3566-08		1	2.53	100	7.89	2.34	0.007	0.008	0.001	0.001	06/14/2017	18:03
11	I3566-10		1	2.51	100	7.90	1.71	0.006	0.006	0.000	0.000	06/14/2017	18:04
12	I3566-12		1	2.50	100	7.12	1.80	0.008	0.008	0.000	0.000	06/14/2017	18:04
13	I3566-12DUP		1	2.53	100	7.08	1.54	0.007	0.007	0.000	0.000	06/14/2017	18:05
14	I3566-12MSPre	40	2	2.54	100	7.23	2.48	0.010	0.349	0.339	0.434	06/14/2017	18:05
15	CCV2	0.5	1	100	100		2.35	0.000	0.392	0.392	0.502	06/14/2017	18:06
16	CCB2		1	100	100		2.30	0.000	0.000	0.000	0.000	06/14/2017	18:06
17	I3566-12MS2Ins	1284	40	2.51	100	7.85	1.72	0.011	0.633	0.622	0.797	06/14/2017	18:07
18	I3566-12MS3Post	40	2	2.52	100	7.77	2.25	0.012	0.370	0.358	0.458	06/14/2017	18:07
19	I3566-14		1	2.55	100	7.33	2.10	0.012	0.012	0.000	0.000	06/14/2017	18:08
20	I3680-01		1	2.51	100	7.41	1.71	0.010	0.011	0.001	0.001	06/14/2017	18:08
21	I3681-01		1	2.54	100	7.56	1.63	0.017	0.019	0.002	0.002	06/14/2017	18:09
22	CCV3	0.5	1	100	100		2.46	0.000	0.393	0.393	0.503	06/14/2017	18:09
23	CCB3		1	100	100		1.66	0.000	0.000	0.000	0.000	06/14/2017	18:10