


7471A

LB88098
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45160



A= 0.0000e+000

B= 2.1888e-004

C= 3.9162e-002

Rho= 0.9995179

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	-41	0.000	-41				
0.2	0.200	0.220	0.020	825	0.0 %	825				
2.5	2.500	2.488	-0.012	11189	0.0 %	11189				
5.0	5.000	4.831	-0.169	21891	0.0 %	21891				
7.5	7.500	7.707	0.207	35032	0.0 %	35032				
10.0	10.000	9.924	-0.076	45160	0.0 %	45160				

LB88098 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 15 Jun 2017 09:06:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-41	-	0.0000	7471A	PPB	15 Jun 2017 09:12:46	S	Std
0.2 - 1	50.2	825	-	0.2000	7471A	PPB	15 Jun 2017 09:15:04	S	Std
2.5 - 1	52.5	11189	-	2.5000	7471A	PPB	15 Jun 2017 09:17:22	S	Std
5.0 - 1	55	21891	-	5.0000	7471A	PPB	15 Jun 2017 09:19:39	S	Std
7.5 - 1	57.5	35032	-	7.5000	7471A	PPB	15 Jun 2017 09:24:30	S	Std
10.0 - 1	510	45160	-	10.0000	7471A	PPB	15 Jun 2017 09:29:22	S	Std
ICV99 - 1	ICV99	17037	3.7683	-	7471A	PPB	15 Jun 2017 09:32:08	U	SMPL
ICB99 - 1	ICB99	-67	0.0245	-	7471A	PPB	15 Jun 2017 09:34:23	U	SMPL
CCV49 - 1	CCV49	22584	4.9824	-	7471A	PPB	15 Jun 2017 09:36:39	U	SMPL
CCB49 - 1	CCB49	-94	0.0186	-	7471A	PPB	15 Jun 2017 09:38:55	U	SMPL
CRA - 1	CRA	858	0.2270	-	7471A	PPB	15 Jun 2017 09:41:10	U	SMPL
HighStd - 1	HighStd	45953	10.0975	-	7471A	PPB	15 Jun 2017 09:43:26	U	SMPL
ChkStd - 1	ChkStd	30957	6.8152	-	7471A	PPB	15 Jun 2017 09:45:42	U	SMPL
PB99794BL - 1	PBS	-87	0.0201	-	7471A	PPB	15 Jun 2017 09:47:59	U	SMPL
PB99794BS - 1	LCSS	16050	3.5522	-	7471A	PPB	15 Jun 2017 09:50:16	U	SMPL
I3660-01 - 1	RO-132	949	0.2469	-	7471A	PPB	15 Jun 2017 09:52:33	U	SMPL
I3660-01DUP - 1	RO-132DUP	1029	0.2644	-	7471A	PPB	15 Jun 2017 09:54:50	U	SMPL
I3660-01MS - 1	RO-132MS	17359	3.8388	-	7471A	PPB	15 Jun 2017 09:57:08	U	SMPL
I3660-01MSD - 1	RO-132MSD	17788	3.9327	-	7471A	PPB	15 Jun 2017 09:59:26	U	SMPL
I3660-05 - 1	RO-153	465	0.1409	-	7471A	PPB	15 Jun 2017 10:01:42	U	SMPL
CCV50 - 1	CCV50	22574	4.9802	-	7471A	PPB	15 Jun 2017 10:03:58	U	SMPL
CCB50 - 1	CCB50	-62	0.0256	-	7471A	PPB	15 Jun 2017 10:06:14	U	SMPL
I3660-09 - 1	RO-351	10703	2.3819	-	7471A	PPB	15 Jun 2017 10:08:30	U	SMPL
I3660-13 - 1	RO-350	1119	0.2841	-	7471A	PPB	15 Jun 2017 10:10:46	U	SMPL
I3660-17 - 1	RO-72-11941	2578	0.6034	-	7471A	PPB	15 Jun 2017 10:13:02	U	SMPL
I3663-01 - 1	ONSITE-001-A	6477	1.4569	-	7471A	PPB	15 Jun 2017 10:15:19	U	SMPL
I3663-03 - 1	ONSITE-001-B	7230	1.6217	-	7471A	PPB	15 Jun 2017 10:17:35	U	SMPL
I3664-01 - 1	ONSITE-001A	1316	0.3272	-	7471A	PPB	15 Jun 2017 10:19:52	U	SMPL
I3664-03 - 1	ONSITE-001B	1988	0.4743	-	7471A	PPB	15 Jun 2017 10:22:09	U	SMPL
I3665-04 - 1	HA3-WC1	4051	0.9259	-	7471A	PPB	15 Jun 2017 10:24:27	U	SMPL
I3665-09 - 1	HA3-WC2	311	0.1072	-	7471A	PPB	15 Jun 2017 10:26:44	U	SMPL
I3670-01 - 1	TP-CLINTON-AVE-1-2	1389	0.3432	-	7471A	PPB	15 Jun 2017 10:29:00	U	SMPL
CCV51 - 1	CCV51	22685	5.0045	-	7471A	PPB	15 Jun 2017 10:31:17	U	SMPL
CCB51 - 1	CCB51	-53	0.0276	-	7471A	PPB	15 Jun 2017 10:33:33	U	SMPL
I3678-02 - 1	WASTE-B	1016	0.2615	-	7471A	PPB	15 Jun 2017 10:35:49	U	SMPL
I3680-01 - 1	TR-05-061317	1183	0.2981	-	7471A	PPB	15 Jun 2017 10:38:06	U	SMPL
I3681-01 - 1	RO-272031-72-11978-COMP	251	0.0941	-	7471A	PPB	15 Jun 2017 10:40:22	U	SMPL
I3660-01LX5 - 1		155	0.0731	-	7471A	PPB	15 Jun 2017 10:42:38	U	SMPL
I3660-01A - 1		18556	4.1008	-	7471A	PPB	15 Jun 2017 10:44:55	U	SMPL
CCV52 - 1	CCV52	22679	5.0032	-	7471A	PPB	15 Jun 2017 10:47:12	U	SMPL
CCB52 - 1	CCB52	-62	0.0256	-	7471A	PPB	15 Jun 2017 10:49:29	U	SMPL