

LB88101

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

6/15/2017 10:33

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.978	0.0	0.108	
ICB1	0.001	0.0	0.002	
CCV1	0.979	0.0	0.108	
CCB1	0.001	0.0	0.002	
PB99788BL	0.005	0.0	0.003	
PB99788BS	0.983	0.0	0.108	
I3665-04	0.006	0.0	0.003	
I3665-04DUP	0.004	0.0	0.002	
I3665-04MS	0.978	0.0	0.108	
I3665-04MSD	0.972	0.0	0.107	
I3665-09	0.001	0.0	0.002	
CCV2	0.975	0.0	0.107	
CCB2	0.001	0.0	0.002	

N	13
Mean	0.453
SD	0.5059
CV%	111.73

Aquakem v. 7.2AQ1

AK

Results from time period:

Thu Jun 15 09:55:21 2017

Thu Jun 15 10:29:13 2017

Sample Id	Sam/Ctr/c/	Test short n:	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0528	mg/l	6/15/2017 9:55:21	
0.1PPM	A	Phenolics-	P	0.1023	mg/l	6/15/2017 9:55:22	
0.25PPM	A	Phenolics-	P	0.2534	mg/l	6/15/2017 9:55:23	
0.50PPM	A	Phenolics-	P	0.4948	mg/l	6/15/2017 9:55:24	
1.0PPM	A	Phenolics-	P	0.9921	mg/l	6/15/2017 9:55:25	
2.0PPM	A	Phenolics-	P	2.0046	mg/l	6/15/2017 9:55:26	
ICV1	S	Phenolics-	P	0.978	mg/l	6/15/2017 10:25:14	
ICB1	S	Phenolics-	P	0.0009	mg/l	6/15/2017 10:25:15	
CCV1	S	Phenolics-	P	0.9792	mg/l	6/15/2017 10:25:16	
CCB1	S	Phenolics-	P	0.0013	mg/l	6/15/2017 10:25:17	
PB99788BL	S	Phenolics-	P	0.005	mg/l	6/15/2017 10:25:18	
PB99788BS	S	Phenolics-	P	0.9832	mg/l	6/15/2017 10:25:19	
I3665-04	S	Phenolics-	P	0.0057	mg/l	6/15/2017 10:25:20	
I3665-04DUP	S	Phenolics-	P	0.0042	mg/l	6/15/2017 10:25:21	
I3665-04MS	S	Phenolics-	P	0.9784	mg/l	6/15/2017 10:25:22	
I3665-04MSD	S	Phenolics-	P	0.9722	mg/l	6/15/2017 10:25:23	
I3665-09	S	Phenolics-	P	0.0014	mg/l	6/15/2017 10:25:24	
CCV2	S	Phenolics-	P	0.9753	mg/l	6/15/2017 10:25:25	
CCB2	S	Phenolics-	P	0.0011	mg/l	6/15/2017 10:29:12	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

6/15/2017 10:00


Test Phenolics-

Accepted 6/15/2017 10:00

Factor 9.272
Bias 0.002

Coeff. of det. 0.999952

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0528	0.0500	
2	Phenol-2	0.013	0.1023	0.1000	
3	Phenol-2	0.029	0.2534	0.2500	
4	Phenol-2	0.055	0.4948	0.5000	
5	Phenol-2	0.109	0.9921	1.0000	
6	Phenol-2	0.218	2.0046	2.0000	