

LB88104


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38155



A= 0.0000e+000

B= 2.6362e-004

C= 4.7211e-002

Rho= 0.9997932

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.042	0.042	-18	0.000	-18				
0.2	0.200	0.241	0.041	735	0.0 %	735				
2.5	2.500	2.497	-0.003	9294	0.0 %	9294				
5.0	5.000	4.878	-0.122	18323	0.0 %	18323				
7.5	7.500	7.436	-0.064	28027	0.0 %	28027				
10.0	10.000	10.106	0.106	38155	0.0 %	38155				

LB88104 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Jun 2017 11:25:18

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-18	-	0.00007470A	PPB	15 Jun 2017 11:27:23	S	Std
0.2 - 1	50.2	735	-	0.20007470A	PPB	15 Jun 2017 11:29:31	S	Std
2.5 - 1	52.5	9294	-	2.50007470A	PPB	15 Jun 2017 11:31:39	S	Std
5.0 - 1	55	18323	-	5.00007470A	PPB	15 Jun 2017 11:33:47	S	Std
7.5 - 1	57.5	28027	-	7.50007470A	PPB	15 Jun 2017 11:35:55	S	Std
10.0 - 1	510	38155	-	10.00007470A	PPB	15 Jun 2017 11:38:03	S	Std
ICV01 - 1	ICV01	13989	3.7351	-7470A	PPB	15 Jun 2017 11:41:03	U	SMPL
ICB01 - 1	ICB01	-79	0.0264	-7470A	PPB	15 Jun 2017 11:43:09	U	SMPL
CCV53 - 1	CCV53	17875	4.7595	-7470A	PPB	15 Jun 2017 11:45:16	U	SMPL
CCB53 - 1	CCB53	-112	0.0177	-7470A	PPB	15 Jun 2017 11:47:22	U	SMPL
CRA - 1	CRA	654	0.2196	-7470A	PPB	15 Jun 2017 11:49:29	U	SMPL
HighStd - 1	HighStd	36623	9.7019	-7470A	PPB	15 Jun 2017 11:51:36	U	SMPL
ChkStd - 1	ChkStd	25502	6.7702	-7470A	PPB	15 Jun 2017 11:53:43	U	SMPL
PB99787BL - 1	PBW	-76	0.0272	-7470A	PPB	15 Jun 2017 11:55:50	U	SMPL
PB99787BS - 1	LCSW	14272	3.8097	-7470A	PPB	15 Jun 2017 11:57:57	U	SMPL
PB99787TB - 1	PB99737TB	-59	0.0317	-7470A	PPB	15 Jun 2017 12:00:05	U	SMPL
I3635-03 - 1	CL-02-061217-A	-55	0.0327	-7470A	PPB	15 Jun 2017 12:02:13	U	SMPL
I3635-06 - 1	CL-02-061217-B	-65	0.0301	-7470A	PPB	15 Jun 2017 12:04:21	U	SMPL
I3635-09 - 1	CL-02-061217-C	-51	0.0338	-7470A	PPB	15 Jun 2017 12:06:30	U	SMPL
I3635-12 - 1	CL-02-061217-D	-32	0.0388	-7470A	PPB	15 Jun 2017 12:08:37	U	SMPL
CCV54 - 1	CCV54	18227	4.8523	-7470A	PPB	15 Jun 2017 12:10:44	U	SMPL
CCB54 - 1	CCB54	-78	0.0266	-7470A	PPB	15 Jun 2017 12:12:51	U	SMPL
I3635-15 - 1	CL-02-061217-E	-41	0.0364	-7470A	PPB	15 Jun 2017 12:14:58	U	SMPL
I3636-03 - 1	OR-02-061217	-63	0.0306	-7470A	PPB	15 Jun 2017 12:17:04	U	SMPL
I3643-03 - 1	RO-VNJ217	-62	0.0309	-7470A	PPB	15 Jun 2017 12:19:11	U	SMPL
I3644-02 - 1	DRUMS-M47409-COMP	-16	0.0430	-7470A	PPB	15 Jun 2017 12:21:18	U	SMPL
I3660-03 - 1	RO-132	-8	0.0451	-7470A	PPB	15 Jun 2017 12:23:26	U	SMPL
I3660-07 - 1	RO-153	-3	0.0464	-7470A	PPB	15 Jun 2017 12:25:33	U	SMPL
I3660-11 - 1	RO-351	-15	0.0433	-7470A	PPB	15 Jun 2017 12:27:42	U	SMPL
I3660-15 - 1	RO-350	-45	0.0353	-7470A	PPB	15 Jun 2017 12:29:50	U	SMPL
I3660-19 - 1	RO-72-11941	-68	0.0293	-7470A	PPB	15 Jun 2017 12:31:58	U	SMPL
I3665-05 - 1	HA3-WC1	-35	0.0380	-7470A	PPB	15 Jun 2017 12:34:05	U	SMPL
CCV55 - 1	CCV55	18051	4.8059	-7470A	PPB	15 Jun 2017 12:36:12	U	SMPL
CCB55 - 1	CCB55	-83	0.0253	-7470A	PPB	15 Jun 2017 12:38:19	U	SMPL
I3665-10 - 1	HA3-WC2	-12	0.0440	-7470A	PPB	15 Jun 2017 12:40:26	U	SMPL
I3670-04 - 1	TP-CLINTON-AVE-1-2	-9	0.0448	-7470A	PPB	15 Jun 2017 12:42:33	U	SMPL
I3670-04DUP - 1	TP-CLINTON-AVE-1-2DUP	-41	0.0364	-7470A	PPB	15 Jun 2017 12:44:39	U	SMPL
I3670-04MS - 1	TP-CLINTON-AVE-1-2MS	13514	3.6098	-7470A	PPB	15 Jun 2017 12:46:46	U	SMPL
I3670-04MSD - 1	TP-CLINTON-AVE-1-2MSD	13508	3.6083	-7470A	PPB	15 Jun 2017 12:48:53	U	SMPL
I3670-04LX5 - 1		-123	0.0148	-7470A	PPB	15 Jun 2017 12:51:01	U	SMPL
I3670-04A - 1		14143	3.7757	-7470A	PPB	15 Jun 2017 12:53:09	U	SMPL
CCV56 - 1	CCV56	17955	4.7806	-7470A	PPB	15 Jun 2017 12:55:18	U	SMPL
CCB56 - 1	CCB56	-90	0.0235	-7470A	PPB	15 Jun 2017 12:57:26	U	SMPL