



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88117

Temp.: 22.1 °C

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/15/2017	13:40
2	CAL2	1	Water			7.00	06/15/2017	13:41
3	CAL3	1	Water			10.02	06/15/2017	13:42
4	ICV	1	Water			7.01	06/15/2017	13:45
5	CCV1	1	Water			2.01	06/15/2017	13:46
6	I3725-01	1	Water			7.20	06/15/2017	13:47
7	I3725-01DUP	1	Water			7.21	06/15/2017	13:49
8	CCV2	1	Water			12.01	06/15/2017	13:50

1888117

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3725

WorkList ID : 99629

Date : 6/15/2017 1:28:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2017	Water	I3725-01	pH	Cool 4 deg C	PSEG01	M41	A-598-FRAC-TANK	06/15/2017	9045C

Date/Time 06.15.17 12:28 PM
 Received by: JP
 Relinquished by: JP

Date/Time 06.15.17 3:35 PM
 Received by: JP
 Relinquished by: JP