



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88119

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/15/2017	15:00
2	CAL2	1	Water			7.00	06/15/2017	15:01
3	CAL3	1	Water			10.04	06/15/2017	15:03
4	ICV	1	Water			7.00	06/15/2017	15:05
5	CCV1	1	Water			2.01	06/15/2017	15:06
6	I3698-01	1	Solid	20.02	20	5.80	06/15/2017	15:07
7	I3698-01DUP	1	Solid	20.01	20	5.81	06/15/2017	15:10
8	CCV2	1	Water			12.01	06/15/2017	15:11

2388119

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 13698

WorkList ID : 99626

Date : 6/15/2017 12:39:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3698-01-18	pH	Cool 4 deg C	TULL02	M23	FK-PS-01-007	06/14/2017	9045C

Date/Time 06-13-17 12:40 PM
Received by: JP
Relinquished by: JP

Date/Time 06-13-17 12:40 PM
Received by: JP
Relinquished by: JP