



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88120

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/15/2017	15:30
2	CAL2	1	Water			7.01	06/15/2017	15:31
3	CAL3	1	Water			10.03	06/15/2017	15:33
4	ICV	1	Water			6.99	06/15/2017	15:34
5	CCV1	1	Water			2.01	06/15/2017	15:35
6	I3716-01	1	Solid	20.03	20	8.61	06/15/2017	15:36
7	I3716-01DUP	1	Solid	20.04	20	8.62	06/15/2017	15:37
8	I3716-02	1	Solid	20.02	20	8.70	06/15/2017	15:39
9	I3716-03	1	Solid	20.03	20	8.93	06/15/2017	15:40
10	I3716-04	1	Solid	20.00	20	9.16	06/15/2017	15:42
11	I3716-05	1	Solid	20.03	20	7.63	06/15/2017	15:43
12	I3716-06	1	Solid	20.03	20	8.73	06/15/2017	15:45
13	I3716-07	1	Solid	20.05	20	7.50	06/15/2017	15:47
14	I3716-08	1	Solid	20.03	20	6.55	06/15/2017	15:49
15	CCV2	1	Water			12.02	06/15/2017	15:50

WORKLIST(Hardcopy Internal Chain)

7888120

WorkList Name : PH I3716

WorkList ID : 99633

Date : 6/15/2017 3:43:29 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3716-01	pH	Cool 4 deg C	CHUG01	M32	G2-01	06/15/2017	9045C
	Solid	I3716-02	pH	Cool 4 deg C	CHUG01	M32	G2-02	06/15/2017	9045C
	Solid	I3716-03	pH	Cool 4 deg C	CHUG01	M32	G2-03	06/15/2017	9045C
	Solid	I3716-04	pH	Cool 4 deg C	CHUG01	M32	G2-04	06/15/2017	9045C
	Solid	I3716-05	pH	Cool 4 deg C	CHUG01	M32	G2-05	06/15/2017	9045C
	Solid	I3716-06	pH	Cool 4 deg C	CHUG01	M32	G2-06	06/15/2017	9045C
	Solid	I3716-07	pH	Cool 4 deg C	CHUG01	M32	G2-07	06/15/2017	9045C
	Solid	I3716-08	pH	Cool 4 deg C	CHUG01	M32	G2-08	06/15/2017	9045C

Date/Time

Received by:

Relinquished by:

06-15-17 4:00 PM

JP

JP

Date/Time

Received by:

Relinquished by:

06-15-17 5:32 PM

JP

JP