



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88124

Temp.: 23.7 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/15/2017	16:30
2	CAL2	1	Water			7.00	06/15/2017	16:31
3	CAL3	1	Water			10.03	06/15/2017	16:32
4	ICV	1	Water			6.99	06/15/2017	16:33
5	CCV1	1	Water			2.02	06/15/2017	16:35
6	I3693-03	1	Solid	20.02	20	10.73	06/15/2017	16:36
7	I3693-03DUP	1	Solid	20.03	20	10.74	06/15/2017	16:37
8	I3693-06	1	Solid	20.02	20	10.77	06/15/2017	16:39
9	I3693-09	1	Solid	20.05	20	10.04	06/15/2017	16:40
10	I3693-12	1	Solid	20.01	20	10.25	06/15/2017	16:41
11	I3694-03	1	Solid	20.03	20	7.90	06/15/2017	16:43
12	I3694-06	1	Solid	20.01	20	8.50	06/15/2017	16:45
13	I3694-09	1	Solid	20.03	20	7.96	06/15/2017	16:47
14	I3695-03	1	Solid	20.02	20	5.51	06/15/2017	16:49
15	I3696-03	1	Solid	20.04	20	6.54	06/14/2017	16:50
16	CCV2	1	Water			12.02	06/15/2017	16:51

WORKLIST(Hardcopy Internal Chain)

2B 88124

WorkList Name : CORROSIONITY I3693

WorkList ID : 99618

Date : 6/15/2017 11:35:50 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2017	Solid	I3693-03	Corrosivity	Cool 4 deg C	PSEG05	M23	PL-01-061417-A	06/14/2017	9045C
06/19/2017	Solid	I3693-06	Corrosivity	Cool 4 deg C	PSEG05	M23	PL-01-061417-B	06/14/2017	9045C
06/19/2017	Solid	I3693-09	Corrosivity	Cool 4 deg C	PSEG05	M23	PL-01-061417-C	06/14/2017	9045C
06/19/2017	Solid	I3693-12	Corrosivity	Cool 4 deg C	PSEG05	M23	PL-01-061417-D	06/14/2017	9045C
06/19/2017	Solid	I3694-03	Corrosivity	Cool 4 deg C	PSEG05	M31	NB-08-061417-A	06/14/2017	9045C
06/19/2017	Solid	I3694-06	Corrosivity	Cool 4 deg C	PSEG05	M31	NB-08-061417-B	06/14/2017	9045C
06/19/2017	Solid	I3694-09	Corrosivity	Cool 4 deg C	PSEG05	M31	NB-08-061417-C	06/14/2017	9045C
06/19/2017	Solid	I3695-03	Corrosivity	Cool 4 deg C	PSEG05	M31	BU-02-061417	06/14/2017	9045C
06/19/2017	Solid	I3696-03	Corrosivity	Cool 4 deg C	PSEG01	M31	NB-RO-72-11935	06/14/2017	9045C

Date/Time 06.15.17 11:42 AM

Received by: JN

Relinquished by: _____

Date/Time 06.15.17 6:32 PM

Received by: CP

Relinquished by: JN