

7471A


LB88128

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44994



A= 0.0000e+000

B= 2.2121e-004

C= 1.5478e-002

Rho= 0.9999712

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	-18	0.000	-18				
0.2	0.200	0.200	-0.000	833	0.0 %	833				
2.5	2.500	2.463	-0.037	11065	0.0 %	11065				
5.0	5.000	5.010	0.010	22580	0.0 %	22580				
7.5	7.500	7.546	0.046	34043	0.0 %	34043				
10.0	10.000	9.969	-0.031	44994	0.0 %	44994				

LB88128 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 16 Jun 2017 08:39:46

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-18	-	0.00007471A	PPB	16 Jun 2017 08:52:21	S	Std
0.2 - 1	50.2	833	-	0.20007471A	PPB	16 Jun 2017 08:54:39	S	Std
2.5 - 1	52.5	11065	-	2.50007471A	PPB	16 Jun 2017 08:56:56	S	Std
5.0 - 1	55	22580	-	5.00007471A	PPB	16 Jun 2017 08:59:14	S	Std
7.5 - 1	57.5	34043	-	7.50007471A	PPB	16 Jun 2017 09:01:31	S	Std
10.0 - 1	510	44994	-	10.00007471A	PPB	16 Jun 2017 09:03:47	S	Std
ICV02 - 1	ICV02	16853	3.7436	-7471A	PPB	16 Jun 2017 09:06:29	U	SMPL
ICB02 - 1	ICB02	-97	-0.0060	-7471A	PPB	16 Jun 2017 09:08:45	U	SMPL
CCV57 - 1	CCV57	21944	4.8698	-7471A	PPB	16 Jun 2017 09:11:01	U	SMPL
CCB57 - 1	CCB57	-73	-0.0007	-7471A	PPB	16 Jun 2017 09:13:16	U	SMPL
CRA - 1	CRA	861	0.2059	-7471A	PPB	16 Jun 2017 09:15:32	U	SMPL
HighStd - 1	HighStd	44637	9.8898	-7471A	PPB	16 Jun 2017 09:17:48	U	SMPL
ChkStd - 1	ChkStd	31074	6.8895	-7471A	PPB	16 Jun 2017 09:20:04	U	SMPL
PB99816BL - 1	PBS	-43	0.0060	-7471A	PPB	16 Jun 2017 09:22:21	U	SMPL
PB99816BS - 1	LCSS	14880	3.3071	-7471A	PPB	16 Jun 2017 09:24:37	U	SMPL
I3686-01 - 1	E2-S5-NSW	3757	0.8466	-7471A	PPB	16 Jun 2017 09:26:54	U	SMPL
I3686-02 - 1	E2-S5-NSW(SOIL-DUP)	7080	1.5817	-7471A	PPB	16 Jun 2017 09:29:11	U	SMPL
I3686-03 - 1	E2-S5-WSW	22807	5.0607	-7471A	PPB	16 Jun 2017 09:31:29	U	SMPL
I3686-04 - 1	E2-R5-BASE	70938	15.7080	-7471A	PPB	16 Jun 2017 09:33:46	U	SMPL
I3689-01 - 1	BAYWAY-SWITCH-A	3047	0.6895	-7471A	PPB	16 Jun 2017 09:36:02	U	SMPL
CCV58 - 1	CCV58	21956	4.8725	-7471A	PPB	16 Jun 2017 09:38:19	U	SMPL
CCB58 - 1	CCB58	5	0.0166	-7471A	PPB	16 Jun 2017 09:40:35	U	SMPL
I3689-08 - 1	BAYWAY-SWITCH-B	5086	1.1406	-7471A	PPB	16 Jun 2017 09:42:51	U	SMPL
I3693-01 - 1	PL-01-061417-A	2341	0.5333	-7471A	PPB	16 Jun 2017 09:45:06	U	SMPL
I3693-04 - 1	PL-01-061417-B	731	0.1772	-7471A	PPB	16 Jun 2017 09:47:22	U	SMPL
I3693-07 - 1	PL-01-061417-C	1699	0.3913	-7471A	PPB	16 Jun 2017 09:49:38	U	SMPL
I3693-10 - 1	PL-01-061417-D	1444	0.3349	-7471A	PPB	16 Jun 2017 09:51:55	U	SMPL
I3694-01 - 1	NB-08-061417-A	2355	0.5364	-7471A	PPB	16 Jun 2017 09:54:12	U	SMPL
I3694-04 - 1	NB-08-061417-B	2871	0.6506	-7471A	PPB	16 Jun 2017 09:56:30	U	SMPL
I3694-07 - 1	NB-08-061417-C	1820	0.4181	-7471A	PPB	16 Jun 2017 09:58:47	U	SMPL
I3695-01 - 1	BU-02-061417	1549	0.3581	-7471A	PPB	16 Jun 2017 10:01:05	U	SMPL
I3696-01 - 1	NB-RO-72-11935	881	0.2104	-7471A	PPB	16 Jun 2017 10:03:21	U	SMPL
CCV59 - 1	CCV59	21855	4.8501	-7471A	PPB	16 Jun 2017 10:05:37	U	SMPL
CCB59 - 1	CCB59	-61	0.0020	-7471A	PPB	16 Jun 2017 10:07:54	U	SMPL
I3697-01 - 1	FK-SF-02-001-A	5515	1.2355	-7471A	PPB	16 Jun 2017 10:10:10	U	SMPL
I3697-03 - 1	FK-SF-02-001-B	4427	0.9948	-7471A	PPB	16 Jun 2017 10:12:26	U	SMPL
I3697-05 - 1	FK-SF-02-001-C	4323	0.9718	-7471A	PPB	16 Jun 2017 10:14:43	U	SMPL
I3698-01 - 1	FK-PS-01-007	405	0.1051	-7471A	PPB	16 Jun 2017 10:16:59	U	SMPL
I3698-01DUP - 1	FK-PS-01-007DUP	547	0.1365	-7471A	PPB	16 Jun 2017 10:25:06	U	SMPL
I3698-01MS - 1	FK-PS-01-007MS	15989	3.5525	-7471A	PPB	16 Jun 2017 10:27:22	U	SMPL
I3698-01MSD - 1	FK-PS-01-007MSD	16145	3.5870	-7471A	PPB	16 Jun 2017 10:29:39	U	SMPL
I3724-01 - 1	RELIABLE-WOOD-RCA	652	0.1597	-7471A	PPB	16 Jun 2017 10:31:57	U	SMPL
PB99817BL - 1	PBS	-9	0.0135	-7471A	PPB	16 Jun 2017 10:37:23	U	SMPL
PB99817BS - 1	LCSS	15070	3.3492	-7471A	PPB	16 Jun 2017 10:39:39	U	SMPL
CCV60 - 1	CCV60	22097	4.9036	-7471A	PPB	16 Jun 2017 10:41:55	U	SMPL
CCB60 - 1	CCB60	-28	0.0093	-7471A	PPB	16 Jun 2017 10:44:12	U	SMPL
I3684-01 - 1	ROCKY-ROAD-COMP	197	0.0591	-7471A	PPB	16 Jun 2017 10:46:28	U	SMPL
I3684-05 - 1	SOIL-ROAD-COMP	1012	0.2393	-7471A	PPB	16 Jun 2017 10:48:44	U	SMPL
I3715-01 - 1	WC061417-1	113	0.0405	-7471A	PPB	16 Jun 2017 10:51:01	U	SMPL
I3715-01DUP - 1	WC061417-1DUP	75	0.0321	-7471A	PPB	16 Jun 2017 10:53:17	U	SMPL
I3715-01MS - 1	WC061417-1MS	15632	3.4735	-7471A	PPB	16 Jun 2017 10:55:34	U	SMPL
I3715-01MSD - 1	WC061417-1MSD	15416	3.4257	-7471A	PPB	16 Jun 2017 10:57:51	U	SMPL
I3715-08 - 1	WC061417-2	819	0.1967	-7471A	PPB	16 Jun 2017 11:00:08	U	SMPL
I3732-01 - 1	SB-24COMP	594	0.1469	-7471A	PPB	16 Jun 2017 11:02:25	U	SMPL
I3732-02 - 1	SB-27COMP	2414	0.5495	-7471A	PPB	16 Jun 2017 11:04:43	U	SMPL
PB99827BL - 1	PBS	2	0.0159	-7471A	PPB	16 Jun 2017 11:06:59	U	SMPL
CCV61 - 1	CCV61	22485	4.9895	-7471A	PPB	16 Jun 2017 11:09:15	U	SMPL
CCB61 - 1	CCB61	-48	0.0049	-7471A	PPB	16 Jun 2017 11:11:32	U	SMPL
PB99827BS - 1	LCSS	15857	3.5233	-7471A	PPB	16 Jun 2017 11:13:48	U	SMPL
I3688-01 - 1	G23-0-1.5	2211	0.5046	-7471A	PPB	16 Jun 2017 11:16:05	U	SMPL
I3688-02 - 1	G23-1.5-3	5930	1.3273	-7471A	PPB	16 Jun 2017 11:18:21	U	SMPL
I3688-03 - 1	G17-0-1.5	2492	0.5667	-7471A	PPB	16 Jun 2017 11:20:37	U	SMPL
I3688-04 - 1	G17-1.5-3	8653	1.9296	-7471A	PPB	16 Jun 2017 11:22:54	U	SMPL
I3688-05 - 1	G20-0-1.5	3160	0.7145	-7471A	PPB	16 Jun 2017 11:25:10	U	SMPL
I3688-05DUP - 1	G20-0-1.5DUP	2912	0.6597	-7471A	PPB	16 Jun 2017 11:27:28	U	SMPL
I3688-05MS - 1	G20-0-1.5MS	18463	4.0998	-7471A	PPB	16 Jun 2017 11:29:45	U	SMPL
I3688-05MSD - 1	G20-0-1.5MSD	17990	3.9951	-7471A	PPB	16 Jun 2017 11:32:03	U	SMPL
I3688-06 - 1	G20-1.5-3	6276	1.4038	-7471A	PPB	16 Jun 2017 11:34:20	U	SMPL
CCV62 - 1	CCV62	22415	4.9740	-7471A	PPB	16 Jun 2017 11:36:37	U	SMPL
CCB62 - 1	CCB62	-26	0.0097	-7471A	PPB	16 Jun 2017 11:38:53	U	SMPL
I3688-07 - 1	G16-0-1.5	9819	2.1876	-7471A	PPB	16 Jun 2017 11:41:10	U	SMPL
I3688-08 - 1	G16-1.5-3	7710	1.7210	-7471A	PPB	16 Jun 2017 11:43:27	U	SMPL
I3688-09 - 1	G15-1.5-3	10808	2.4064	-7471A	PPB	16 Jun 2017 11:45:43	U	SMPL
I3688-10 - 1	G25-0-1.5	6064	1.3569	-7471A	PPB	16 Jun 2017 11:48:00	U	SMPL
I3688-11 - 1	G25-1.5-3	6754	1.5096	-7471A	PPB	16 Jun 2017 11:50:17	U	SMPL
I3688-12 - 1	G15-0-1.5	6425	1.4368	-7471A	PPB	16 Jun 2017 11:52:34	U	SMPL
I3688-13 - 1	G13-0-1.5	2302	0.5247	-7471A	PPB	16 Jun 2017 11:54:52	U	SMPL
I3688-14 - 1	G13-1.5-3	7874	1.7573	-7471A	PPB	16 Jun 2017 11:57:09	U	SMPL
I3688-15 - 1	G12-0-1.5	5706	1.2777	-7471A	PPB	16 Jun 2017 11:59:27	U	SMPL
I3688-16 - 1	G12-1.5-3	3893	0.8767	-7471A	PPB	16 Jun 2017 12:01:44	U	SMPL
CCV63 - 1	CCV63	22483	4.9890	-7471A	PPB	16 Jun 2017 12:04:01	U	SMPL

LB88128 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 16 Jun 2017 08:39:46

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
CCB63 - 1	CCB63	-17	0.0117		-7471A	PPB	16 Jun 2017 12:06:18	U	SMPL
I3688-17 - 1	G11-0-1.5	4506	1.0123		-7471A	PPB	16 Jun 2017 12:08:35	U	SMPL
I3688-18 - 1	G11-1.5-3	8691	1.9380		-7471A	PPB	16 Jun 2017 12:10:52	U	SMPL
I3698-01LX5 - 1		82	0.0336		-7471A	PPB	16 Jun 2017 12:13:09	U	SMPL
I3698-01A - 1		17635	3.9166		-7471A	PPB	16 Jun 2017 12:15:26	U	SMPL
I3715-01LX5 - 1		-5	0.0144		-7471A	PPB	16 Jun 2017 12:17:43	U	SMPL
I3715-01A - 1		17281	3.8383		-7471A	PPB	16 Jun 2017 12:20:00	U	SMPL
I3688-05LX5 - 1		587	0.1453		-7471A	PPB	16 Jun 2017 12:22:17	U	SMPL
I3688-05A - 1		20567	4.5652		-7471A	PPB	16 Jun 2017 12:24:35	U	SMPL
I3686-04DLX5 - 1		14305	3.1799		-7471A	PPB	16 Jun 2017 12:26:53	U	SMPL
CCV64 - 1	CCV64	22624	5.0202		-7471A	PPB	16 Jun 2017 12:29:10	U	SMPL
CCB64 - 1	CCB64	-33	0.0082		-7471A	PPB	16 Jun 2017 12:31:27	U	SMPL