



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB88132
Analysis Date: 06/16/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 06/16/2017
Extraction IN Time: 10:20
Extraction OUT Time: 11:05

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB88132BL	LB88132BL	WATER	2	1000	100	3.5147	3.5147	3.02	3.5147	3.5147	0	0
2	LB88132BS	LB88132BS	WATER	2	1000	100	3.0602	3.0602	3.01	3.0792	3.0792	0.019	19
3	LB88132BSD	LB88132BSD	WATER	2	1000	100	3.1529	3.1529	3.05	3.1721	3.1721	0.0192	19.2
4	I3644-01	DRUMS-M47409-COMP	WATER	2	250	100	3.0824	3.0824	3.04	3.0847	3.0847	0.0023	9.2
5	I3661-01	FT-265434-265578-COMP	WATER	2	1000	100	3.0888	3.0888	3.01	3.0891	3.0891	0.0003	0.3
6	I3671-02	GRAB	WATER	2	1000	100	2.9985	2.9985	3.03	3.151	3.151	0.1525	152.5
7	I3691-01	A1013	WATER	2	1000	100	3.0928	3.0928	3.04	3.1008	3.1008	0.008	8
8	I3725-01	A-598-FRAC-TANK	WATER	2	1000	100	3.0671	3.0671	3.05	3.0674	3.0674	0.0003	0.3



QC Batch# LB88132 Test: TPH Analysis Date: 06/16/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2243
pH Paper 0-14	W2186
Sodium Sulfate	EP1754
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP52823
LCSWD	10.00 ML	WP54439
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0019 In OVEN TEMP : 70 Dessicator Time In : 12:00
1.0000 gram Balance: 1.0005 In Time: 11:25
Bal Check Time: 10:35

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 13:30
1.0000 gram Balance: 1.0004 Out Time: 13:00
Bal Check Time: 13:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I3664

WorkList ID : 99686

Date : 6/16/2017 10:09:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2017	Water	I3644-01-17	TPH	Conc H2SO4 4c	PSEG01	K42	DRUMS-M47409-COMP	06/13/2017	1664A
06/16/2017	Water	I3661-01-17	TPH	Conc H2SO4 4c	PSEG01	M62	FT-265434-265578-COMP	06/13/2017	1664A
	Water	I3671-02	TPH	Conc H2SO4 4c	ARAM01	M13	GRAB	06/14/2017	1664A
06/19/2017	Water	I3691-01	TPH	Conc H2SO4 4c	PSEG01	M21	A1013	06/14/2017	1664A
06/20/2017	Water	I3725-01-17	TPH	Conc H2SO4 4c	PSEG01	M41	A-598-FRAC-TANK	06/15/2017	1664A

Date/Time 06-16-17 10:09 AM
 Received by: JD
 Relinquished by: CD

Date/Time 06-16-17 12:33 PM
 Received by: CD
 Relinquished by: JD