



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88144

Temp.: 21.3 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/16/2017	16:00
2	CAL2	1	Water			7.01	06/16/2017	16:01
3	CAL3	1	Water			10.02	06/16/2017	16:03
4	ICV	1	Water			7.00	06/16/2017	16:05
5	CCV1	1	Water			2.01	06/16/2017	16:07
6	I3749-01	1	Solid	20.03	20	7.64	06/16/2017	16:09
7	I3749-01DUP	1	Solid	20.05	20	7.65	06/16/2017	16:10
8	I3749-02	1	Solid	20.02	20	7.60	06/16/2017	16:11
9	CCV2	1	Water			12.01	06/16/2017	16:12

WORKLIST(Hardcopy Internal Chain)

88144

WorkList Name : ph i3749

WorkList ID : 99703

Date : 6/16/2017 2:35:21 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3749-01	pH	Cool 4 deg C	ENTA05	M41	TILCON-MH-001	06/16/2017	9045C
	Solid	I3749-02	pH	Cool 4 deg C	ENTA05	M41	TILCON-PL-001	06/16/2017	9045C

Date/Time 06-17-2017 2:35 PM
 Received by: JP
 Relinquished by: JP

Date/Time 06-17-2017 4:30 PM
 Received by: JP
 Relinquished by: JP