



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88163

Temp.: 21.2 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/17/2017	11:15
2	CAL2	1	Water			7.01	06/17/2017	11:16
3	CAL3	1	Water			10.04	06/17/2017	11:19
4	ICV	1	Water			6.99	06/17/2017	11:20
5	CCV1	1	Water			2.02	06/17/2017	11:23
6	I3688-19	1	Solid	20.02	20	10.81	06/17/2017	11:25
7	I3688-19DUP	1	Solid	20.04	20	10.82	06/17/2017	11:26
8	I3688-20	1	Solid	20.01	20	8.90	06/17/2017	11:28
9	I3688-21	1	Solid	20.02	20	9.24	06/17/2017	11:30
10	I3688-22	1	Solid	20.03	20	8.44	06/17/2017	11:31
11	I3688-23	1	Solid	20.02	20	10.98	06/17/2017	11:33
12	I3688-24	1	Solid	20.02	20	9.31	06/17/2017	11:36
13	I3688-25	1	Solid	20.02	20	8.75	06/17/2017	11:39
14	I3688-26	1	Solid	20.03	20	7.34	06/17/2017	11:42
15	I3688-27	1	Solid	20.00	20	7.70	06/17/2017	11:44
16	CCV2	1	Water			12.02	06/17/2017	11:46
17	I3688-28	1	Solid	20.03	20	9.37	06/17/2017	11:47
18	I3688-29	1	Solid	20.02	20	8.58	06/17/2017	11:49
19	I3688-30	1	Solid	20.03	20	8.63	06/17/2017	11:51
20	I3688-31	1	Solid	20.04	20	7.94	06/17/2017	11:53

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I3688-32	1	Solid	20.02	20	8.31	06/17/2017	11:55
22	I3688-33	1	Solid	20.04	20	8.41	06/17/2017	12:00
23	I3688-34	1	Solid	20.03	20	7.63	06/17/2017	12:03
24	I3688-35	1	Solid	20.03	20	8.40	06/17/2017	12:05
25	I3688-36	1	Solid	20.01	20	7.22	06/17/2017	12:07
26	I3688-36DUP	1	Solid	20.03	20	7.23	06/17/2017	12:09
27	CCV3	1	Water			12.01	06/17/2017	12:10

20588163

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2688

WorkList ID : 99715

Date : 6/17/2017 8:11:14 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3688-19	Corrosivity	Cool 4 deg C	TECT01	M31	G23-0-1.5	06/14/2017	9045C
	Solid	I3688-20	Corrosivity	Cool 4 deg C	TECT01	M31	G23-1.5-3	06/14/2017	9045C
	Solid	I3688-21	Corrosivity	Cool 4 deg C	TECT01	M31	G17-0-1.5	06/14/2017	9045C
	Solid	I3688-22	Corrosivity	Cool 4 deg C	TECT01	M31	G17-1.5-3	06/14/2017	9045C
	Solid	I3688-23	Corrosivity	Cool 4 deg C	TECT01	M31	G20-0-1.5	06/14/2017	9045C
	Solid	I3688-24	Corrosivity	Cool 4 deg C	TECT01	M31	G20-1.5-3	06/14/2017	9045C
	Solid	I3688-25	Corrosivity	Cool 4 deg C	TECT01	M31	G16-0-1.5	06/14/2017	9045C
	Solid	I3688-26	Corrosivity	Cool 4 deg C	TECT01	M31	G16-1.5-3	06/14/2017	9045C
	Solid	I3688-27	Corrosivity	Cool 4 deg C	TECT01	M31	G15-1.5-3	06/14/2017	9045C
	Solid	I3688-28	Corrosivity	Cool 4 deg C	TECT01	M31	G25-0-1.5	06/14/2017	9045C
	Solid	I3688-29	Corrosivity	Cool 4 deg C	TECT01	M31	G25-1.5-3	06/14/2017	9045C
	Solid	I3688-30	Corrosivity	Cool 4 deg C	TECT01	M31	G15-0-1.5	06/14/2017	9045C
	Solid	I3688-31	Corrosivity	Cool 4 deg C	TECT01	M31	G13-0-1.5	06/14/2017	9045C
	Solid	I3688-32	Corrosivity	Cool 4 deg C	TECT01	M31	G13-1.5-3	06/14/2017	9045C
	Solid	I3688-33	Corrosivity	Cool 4 deg C	TECT01	M31	G12-0-1.5	06/14/2017	9045C
	Solid	I3688-34	Corrosivity	Cool 4 deg C	TECT01	M31	G12-1.5-3	06/14/2017	9045C
	Solid	I3688-35	Corrosivity	Cool 4 deg C	TECT01	M31	G11-0-1.5	06/14/2017	9045C
	Solid	I3688-36	Corrosivity	Cool 4 deg C	TECT01	M31	G11-1.5-3	06/14/2017	9045C

Date/Time 06-17-17 10:00 AM
Received by: JP
Relinquished by: CS

Date/Time 06-17-17 4:00 PM
Received by: JP
Relinquished by: JP