



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88166

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	06/18/2017	12:15
2	CAL2	1	Water			7.01	06/18/2017	12:16
3	CAL3	1	Water			10.03	06/18/2017	12:17
4	ICV	1	Water			7.01	06/18/2017	12:19
5	CCV1	1	Water			2.01	06/18/2017	12:20
6	I3736-13	1	Solid	20.03	20	9.97	06/18/2017	12:23
7	I3736-13DUP	1	Solid	20.05	20	9.98	06/18/2017	12:25
8	I3736-14	1	Solid	20.02	20	10.05	06/18/2017	12:26
9	I3736-15	1	Solid	20.03	20	10.53	06/18/2017	12:27
10	I3736-16	1	Solid	20.01	20	8.83	06/18/2017	12:30
11	I3736-17	1	Solid	20.04	20	10.33	06/18/2017	12:31
12	I3736-18	1	Solid	20.05	20	8.53	06/18/2017	12:33
13	I3736-19	1	Solid	20.02	20	8.71	06/18/2017	12:36
14	I3736-20	1	Solid	20.04	20	7.81	06/18/2017	12:38
15	I3736-21	1	Solid	20.02	20	10.13	06/18/2017	12:40
16	CCV2	1	Water			12.01	06/18/2017	12:42
17	I3736-22	1	Solid	20.02	20	8.65	06/18/2017	12:44
18	I3736-23	1	Solid	20.03	20	9.47	06/18/2017	12:46
19	I3736-24	1	Solid	20.01	20	8.38	06/18/2017	12:49
20	CCV3	1	Water			12.01	06/18/2017	12:50

WORKLIST(Hardcopy Internal Chain)

203 88166

WorkList Name : CORROSIONITY I3736

WorkList ID : 99716

Date : 6/18/2017 11:25:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3736-13	Corrosivity	Cool 4 deg C	TECT01	M52	G24-0-1.5	06/15/2017	9045C
	Solid	I3736-14	Corrosivity	Cool 4 deg C	TECT01	M52	G24-1.5-3	06/15/2017	9045C
	Solid	I3736-15	Corrosivity	Cool 4 deg C	TECT01	M52	G22-0-1.5	06/15/2017	9045C
	Solid	I3736-16	Corrosivity	Cool 4 deg C	TECT01	M52	G22-1.5-3	06/15/2017	9045C
	Solid	I3736-17	Corrosivity	Cool 4 deg C	TECT01	M52	G19-0-1.5	06/15/2017	9045C
	Solid	I3736-18	Corrosivity	Cool 4 deg C	TECT01	M52	G19-1.5-3	06/15/2017	9045C
	Solid	I3736-19	Corrosivity	Cool 4 deg C	TECT01	M52	G14-0-1.5	06/15/2017	9045C
	Solid	I3736-20	Corrosivity	Cool 4 deg C	TECT01	M52	G14-1.5-3	06/15/2017	9045C
	Solid	I3736-21	Corrosivity	Cool 4 deg C	TECT01	M52	G18-0-1.5	06/15/2017	9045C
	Solid	I3736-22	Corrosivity	Cool 4 deg C	TECT01	M52	G18-1.5-3	06/15/2017	9045C
	Solid	I3736-23	Corrosivity	Cool 4 deg C	TECT01	M52	G21-0-1.5	06/15/2017	9045C
	Solid	I3736-24	Corrosivity	Cool 4 deg C	TECT01	M52	G21-1.5-3	06/15/2017	9045C

Date/Time 06-18-18 11:25 AM
 Received by: 711
 Relinquished by: [Signature]

Date/Time 06-18-18 5:00 PM
 Received by: [Signature]
 Relinquished by: JD