

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB88167

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP55256
calibration std. phosphate 1 ppm	WP55681
calibration std. phosphate 0.5 ppm	WP55680
calibration std. phosphate 0.3 ppm	WP55679
calibration std. 0 ppm	WP55675
phosphate ICV std.	WP55682
phosphate CCV std.	WP55683
calibration std. phosphate 0.1 ppm	WP55678
calibration std. phosphate 0.05 ppm	WP55677
Combined reagent	WP55719
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
Ascorbic Acid	WP55718
Combined reagent	WP55719

Intercept: -0.0015Slope: 0.6509Regression: 0.999852

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	06/16/2017	11:22
2	CAL2	0.05	1	50	50	0.031	06/16/2017	11:23
3	CAL3	0.10	1	50	50	0.065	06/16/2017	11:23
4	CAL4	0.30	1	50	50	0.186	06/16/2017	11:24
5	CAL5	0.50	1	50	50	0.329	06/16/2017	11:24
6	CAL6	1.00	1	50	50	0.649	06/16/2017	11:25



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB88167

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.319	0.492	06/16/2017	11:25
2	ICB		1	50	50	0.000	0.002	06/16/2017	11:26
3	CCV1	0.50	1	50	50	0.327	0.505	06/16/2017	11:26
4	CCB1		1	50	50	0.000	0.002	06/16/2017	11:27
5	PB99842BL		1	50	50	0.000	0.002	06/16/2017	11:27
6	PB99842BS	0.50	1	50	50	0.318	0.491	06/16/2017	11:28
7	I3721-01		1	50	50	2.420	3.720	06/16/2017	11:28
8	I3721-01DUP		1	50	50	2.425	3.728	06/16/2017	11:29
9	I3721-01MS	0.50	1	50	50	2.496	3.837	06/16/2017	11:29
10	I3721-01MSD	0.50	1	50	50	2.492	3.831	06/16/2017	11:30
11	I3721-01		500	50	50	0.209	0.323	06/16/2017	11:52
12	I3721-01DUP		500	50	50	0.211	0.326	06/16/2017	11:52
13	I3721-01MS	0.50	500	50	50	0.199	0.308	06/16/2017	11:53
14	I3721-01MSD	0.50	500	50	50	0.203	0.314	06/16/2017	11:53
15	CCV2	0.50	1	50	50	0.328	0.506	06/16/2017	11:54
16	CCB2		1	50	50	0.000	0.002	06/16/2017	11:54