

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 06/19/2017

DATE & TIME IN: 06/19/2017 11:00

OVEN TEMP: 104 °C

Run Number: LB88174

DATE & TIME OUT: 06/19/2017 15:10

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB88174BL	LB88174BL	1.4367	1.4367	100	1.4367	1.4367	1.4367	0	0
2	LB88174BS	LB88174BS	1.3085	1.3085	100	1.3628	1.3628	1.3628	0.0543	543
3	I3760-01	TZBUD-20170616	1.4042	1.4042	300	1.405	1.405	1.405	0.0008	2.7
4	I3760-01DUP	TZBUD-20170616DUP	1.436	1.436	300	1.4369	1.4369	1.4369	0.0009	3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

2088174

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS I3760

WorkList ID : 99750

Date : 6/19/2017 10:19:49 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3760-01-E	TSS	Cool 4 deg C	HDR103	N22	TZBUD-20170616	06/16/2017	SM2540 D

Date/Time 06-19-17 10:20 AM
Received by: JD
Relinquished by: CF

Date/Time 06-19-17 12:00 PM
Received by: CF
Relinquished by: JD