


7471A

LB88125
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45144



A= 0.0000e+000

B= 2.2348e-004

C= -2.4236e-002

Rho= 0.9998887

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	101	0.000	101				
0.2	0.200	0.211	0.011	1054	0.0 %	1054				
2.5	2.500	2.563	0.063	11579	0.0 %	11579				
5.0	5.000	4.917	-0.083	22110	0.0 %	22110				
7.5	7.500	7.446	-0.054	33425	0.0 %	33425				
10.0	10.000	10.064	0.064	45144	0.0 %	45144				

LB88185 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 19 Jun 2017 14:04:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	101	-	0.0000	7471A	PPB	19 Jun 2017 14:07:11	S	Std
0.2 - 1	SD.2	1054	-	0.2000	7471A	PPB	19 Jun 2017 14:09:29	S	Std
2.5 - 1	SD.5	11579	-	2.5000	7471A	PPB	19 Jun 2017 14:11:47	S	Std
5.0 - 1	SD	22110	-	5.0000	7471A	PPB	19 Jun 2017 14:14:04	S	Std
7.5 - 1	SD.5	33425	-	7.5000	7471A	PPB	19 Jun 2017 14:16:21	S	Std
10.0 - 1	SD	45144	-	10.0000	7471A	PPB	19 Jun 2017 14:18:37	S	Std
ICV05 - 1	ICV05	17645	3.9190	-	7471A	PPB	19 Jun 2017 14:21:18	U	SMPL
ICB05 - 1	ICB05	0	-0.0242	-	7471A	PPB	19 Jun 2017 14:23:33	U	SMPL
CCV77 - 1	CCV77	22818	5.0751	-	7471A	PPB	19 Jun 2017 14:28:22	U	SMPL
CCB77 - 1	CCB77	11	-0.0218	-	7471A	PPB	19 Jun 2017 14:30:37	U	SMPL
CRA - 1	CRA	990	0.1970	-	7471A	PPB	19 Jun 2017 14:32:53	U	SMPL
HighStd - 1	HighStd	45112	10.0573	-	7471A	PPB	19 Jun 2017 14:35:09	U	SMPL
ChkStd - 1	ChkStd	30765	6.8511	-	7471A	PPB	19 Jun 2017 14:37:25	U	SMPL
PB99866BL - 1	PBS	99	-0.0021	-	7471A	PPB	19 Jun 2017 14:39:41	U	SMPL
PB99866BS - 1	LCSS	16096	3.5729	-	7471A	PPB	19 Jun 2017 14:41:57	U	SMPL
I3703-01 - 1	CLEANSAND-1	685	0.1288	-	7471A	PPB	19 Jun 2017 14:44:14	U	SMPL
I3736-01 - 1	G24-0-1.5	3589	0.7778	-	7471A	PPB	19 Jun 2017 14:46:32	U	SMPL
I3736-02 - 1	G24-1.5-3	3753	0.8145	-	7471A	PPB	19 Jun 2017 14:48:49	U	SMPL
I3736-03 - 1	G22-0-1.5	4757	1.0389	-	7471A	PPB	19 Jun 2017 14:51:07	U	SMPL
I3736-04 - 1	G22-1.5-3	8568	1.8905	-	7471A	PPB	19 Jun 2017 14:53:23	U	SMPL
CCV78 - 1	CCV78	21867	4.8626	-	7471A	PPB	19 Jun 2017 14:55:39	U	SMPL
CCB78 - 1	CCB78	12	-0.0216	-	7471A	PPB	19 Jun 2017 14:57:55	U	SMPL
I3736-05 - 1	G19-0-1.5	7453	1.6414	-	7471A	PPB	19 Jun 2017 15:00:11	U	SMPL
I3736-06 - 1	G19-1.5-3	3326	0.7191	-	7471A	PPB	19 Jun 2017 15:02:28	U	SMPL
I3736-07 - 1	G14-0-1.5	10865	2.4039	-	7471A	PPB	19 Jun 2017 15:04:43	U	SMPL
I3736-08 - 1	G14-1.5-3	6611	1.4532	-	7471A	PPB	19 Jun 2017 15:07:00	U	SMPL
I3736-09 - 1	G18-0-1.5	6365	1.3982	-	7471A	PPB	19 Jun 2017 15:09:16	U	SMPL
I3736-10 - 1	G18-1.5-3	4950	1.0820	-	7471A	PPB	19 Jun 2017 15:11:33	U	SMPL
I3736-11 - 1	G21-0-1.5	9403	2.0771	-	7471A	PPB	19 Jun 2017 15:13:51	U	SMPL
I3736-12 - 1	G21-1.5-3	2855	0.6138	-	7471A	PPB	19 Jun 2017 15:16:08	U	SMPL
I3745-01 - 1	FK-GF-02-006-A	887	0.1740	-	7471A	PPB	19 Jun 2017 15:18:26	U	SMPL
I3745-01DUP - 1	FK-GF-02-006-ADUP	956	0.1894	-	7471A	PPB	19 Jun 2017 15:20:42	U	SMPL
CCV79 - 1	CCV79	22214	4.9401	-	7471A	PPB	19 Jun 2017 15:22:58	U	SMPL
CCB79 - 1	CCB79	15	-0.0209	-	7471A	PPB	19 Jun 2017 15:25:14	U	SMPL
I3745-01MS - 1	FK-GF-02-006-AMS	17022	3.7798	-	7471A	PPB	19 Jun 2017 15:27:31	U	SMPL
I3745-01MSD - 1	FK-GF-02-006-AMSD	17129	3.8037	-	7471A	PPB	19 Jun 2017 15:29:47	U	SMPL
I3745-03 - 1	FK-GF-02-006-B	716	0.1358	-	7471A	PPB	19 Jun 2017 15:32:03	U	SMPL
PB99870BL - 1	PBS	9	-0.0222	-	7471A	PPB	19 Jun 2017 15:37:06	U	SMPL
PB99870BS - 1	LCSS	16054	3.5635	-	7471A	PPB	19 Jun 2017 15:39:22	U	SMPL
I3743-01 - 1	FNDFL-PLUM-061417-COMP	334	0.0504	-	7471A	PPB	19 Jun 2017 15:41:39	U	SMPL
I3743-03 - 1	FNDFL-ESTN-061417-COMP	1149	0.2325	-	7471A	PPB	19 Jun 2017 15:43:57	U	SMPL
I3743-05 - 1	FNDFL-VCTY-061517-COMP	1436	0.2967	-	7471A	PPB	19 Jun 2017 15:46:14	U	SMPL
I3747-01 - 1	E2-R9-BASE-02	4176	0.9090	-	7471A	PPB	19 Jun 2017 15:58:17	U	SMPL
I3747-01DUP - 1	E2-R9-BASE-02DUP	4004	0.8706	-	7471A	PPB	19 Jun 2017 16:00:33	U	SMPL
CCV80 - 1	CCV80	22424	4.9871	-	7471A	PPB	19 Jun 2017 16:02:49	U	SMPL
CCB80 - 1	CCB80	45	-0.0142	-	7471A	PPB	19 Jun 2017 16:05:06	U	SMPL
I3747-02 - 1	I3747-01MS	22694	5.0474	-	7471A	PPB	19 Jun 2017 16:07:22	U	SMPL
I3747-03 - 1	I3747-01MSD	21372	4.7520	-	7471A	PPB	19 Jun 2017 16:09:39	U	SMPL
I3747-04 - 1	E2-R5-WSW	72984	16.2861	-	7471A	PPB	19 Jun 2017 16:11:55	U	SMPL
I3747-05 - 1	E2-R5-NSW	10095	2.2318	-	7471A	PPB	19 Jun 2017 16:14:11	U	SMPL
I3749-01 - 1	TILCON-MH-001	419	0.0694	-	7471A	PPB	19 Jun 2017 16:16:29	U	SMPL
I3749-02 - 1	TILCON-PL-001	395	0.0640	-	7471A	PPB	19 Jun 2017 16:18:46	U	SMPL
I3759-01 - 1	SU-02-061617	1575	0.3277	-	7471A	PPB	19 Jun 2017 16:21:03	U	SMPL
CCV81 - 1	CCV81	21823	4.8527	-	7471A	PPB	19 Jun 2017 16:23:21	U	SMPL
CCB81 - 1	CCB81	-5	-0.0254	-	7471A	PPB	19 Jun 2017 16:25:38	U	SMPL
I3745-01LX5 - 1		212	0.0231	-	7471A	PPB	19 Jun 2017 16:27:55	U	SMPL
I3745-01A - 1		18705	4.1559	-	7471A	PPB	19 Jun 2017 16:30:11	U	SMPL
I3747-01LX5 - 1		766	0.1469	-	7471A	PPB	19 Jun 2017 16:32:27	U	SMPL
I3747-01A - 1		21639	4.8116	-	7471A	PPB	19 Jun 2017 16:34:44	U	SMPL
I3747-04DLX5 - 1		14602	3.2390	-	7471A	PPB	19 Jun 2017 16:37:00	U	SMPL
CCV82 - 1	CCV82	22398	4.9812	-	7471A	PPB	19 Jun 2017 16:39:17	U	SMPL
CCB82 - 1	CCB82	26	-0.0184	-	7471A	PPB	19 Jun 2017 16:41:33	U	SMPL