

245.1


LB88193

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34290



A= 0.0000e+000

B= 2.8832e-004

C= -4.3219e-003

Rho= 0.9997795

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	-23	0.000	-23				
0.2	0.200	0.175	-0.025	622	0.0 %	622				
2.5	2.500	2.457	-0.043	8537	0.0 %	8537				
5.0	5.000	5.088	0.088	17663	0.0 %	17663				
7.5	7.500	7.608	0.108	26404	0.0 %	26404				
10.0	10.000	9.882	-0.118	34290	0.0 %	34290				

LB88193 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 19 Jun 2017 17:16:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-23	-	0.0000	245.1	PPB	19 Jun 2017 17:17:43	S	Std
0.2 - 1	50.2	622	-	0.2000	245.1	PPB	19 Jun 2017 17:19:51	S	Std
2.5 - 1	52.5	8537	-	2.5000	245.1	PPB	19 Jun 2017 17:22:00	S	Std
5.0 - 1	55	17663	-	5.0000	245.1	PPB	19 Jun 2017 17:24:08	S	Std
7.5 - 1	57.5	26404	-	7.5000	245.1	PPB	19 Jun 2017 17:26:15	S	Std
10.0 - 1	510	34290	-	10.0000	245.1	PPB	19 Jun 2017 17:28:22	S	Std
ICV06 - 1	ICV06	13803	3.9754	-245.1		PPB	19 Jun 2017 17:33:59	U	SMPL
ICB06 - 1	ICB06	-50	-0.0187	-245.1		PPB	19 Jun 2017 17:36:05	U	SMPL
CCV83 - 1	CCV83	17289	4.9804	-245.1		PPB	19 Jun 2017 17:38:11	U	SMPL
CCB83 - 1	CCB83	-46	-0.0176	-245.1		PPB	19 Jun 2017 17:40:18	U	SMPL
CRA - 1	CRA	617	0.1736	-245.1		PPB	19 Jun 2017 17:42:24	U	SMPL
HighStd - 1	HighStd	33772	9.7328	-245.1		PPB	19 Jun 2017 17:44:31	U	SMPL
ChkStd - 1	ChkStd	24263	6.9912	-245.1		PPB	19 Jun 2017 17:46:38	U	SMPL
PB99896BL - 1	PBW	62	0.0136	-245.1		PPB	19 Jun 2017 17:48:45	U	SMPL
PB99896BS - 1	LCSW	14272	4.1106	-245.1		PPB	19 Jun 2017 17:50:53	U	SMPL
I3760-01 - 1	TZBUD-20170616	67	0.0150	-245.1		PPB	19 Jun 2017 17:53:00	U	SMPL
I3760-01DUP - 1	TZBUD-20170616DUP	91	0.0219	-245.1		PPB	19 Jun 2017 17:55:08	U	SMPL
I3760-01MS - 1	TZBUD-20170616MS	14139	4.0722	-245.1		PPB	19 Jun 2017 17:57:16	U	SMPL
I3760-01MSD - 1	TZBUD-20170616MSD	14309	4.1212	-245.1		PPB	19 Jun 2017 17:59:25	U	SMPL
CCV84 - 1	CCV84	17113	4.9297	-245.1		PPB	19 Jun 2017 18:01:31	U	SMPL
CCB84 - 1	CCB84	-65	-0.0231	-245.1		PPB	19 Jun 2017 18:03:38	U	SMPL
I3760-01LX5 - 1		-40	-0.0159	-245.1		PPB	19 Jun 2017 18:05:45	U	SMPL
I3760-01A - 1		14296	4.1175	-245.1		PPB	19 Jun 2017 18:11:37	U	SMPL
CCV85 - 1	CCV85	16622	4.7881	-245.1		PPB	19 Jun 2017 18:13:43	U	SMPL
CCB85 - 1	CCB85	-55	-0.0202	-245.1		PPB	19 Jun 2017 18:15:50	U	SMPL