



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88194

Temp.: 24.3 °C

BalanceID: WC SC-4

SLOPE.: 99.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/19/2017	17:00
2	CAL2	1	Water			7.01	06/19/2017	17:01
3	CAL3	1	Water			10.03	06/19/2017	17:03
4	ICV	1	Water			6.99	06/19/2017	17:05
5	CCV1	1	Water			2.01	06/19/2017	17:06
6	I3783-01	1	Solid	20.03	20	8.34	06/19/2017	17:08
7	I3784-01	1	Solid	20.02	20	8.31	06/19/2017	17:09
8	I3784-03	1	Solid	20.04	20	8.20	06/19/2017	17:10
9	I3784-05	1	Solid	20.05	20	7.89	06/19/2017	17:12
10	I3784-05DUP	1	Solid	20.03	20	7.90	06/19/2017	17:15
11	CCV2	1	Water			12.02	06/19/2017	17:16

20388194

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 13783

WorkList ID : 99779

Date : 6/19/2017 4:38:09 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3783-01	pH	Cool 4 deg C	AREC01	M43	RS-1	06/19/2017	9045C
	Solid	I3784-01	pH	Cool 4 deg C	UNIO02	M42	P.56-A-COMP	06/19/2017	9045C
	Solid	I3784-03	pH	Cool 4 deg C	UNIO02	M42	P.56-B-COMP	06/19/2017	9045C
	Solid	I3784-05	pH	Cool 4 deg C	UNIO02	M42	P.56-C-COMP	06/19/2017	9045C

Date/Time 06-19-17 4:40 PM
Received by: JD
Relinquished by: JD

Date/Time 06-19-17 7:00 PM
Received by: JD
Relinquished by: JD