



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB88202
Analysis Date: 06/20/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 06/20/2017
Extraction IN Time: 08:35
Extraction OUT Time: 09:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB88202BL	LB88202BL	WATER	2	1000	100	3.1496	3.1496	3.03	3.1496	3.1496	0	0
2	LB88202BS	LB88202BS	WATER	2	1000	100	3.0657	3.0657	3.01	3.0848	3.0848	0.0191	19.1
3	LB88202BSD	LB88202BSD	WATER	2	1000	100	3.0388	3.0388	3.02	3.0581	3.0581	0.0193	19.3
4	I3738-02	SEWER-OUTLET-GRAB	WATER	2	1000	100	3.0967	3.0967	3.05	3.097	3.097	0.0003	0.3



QC Batch# LB88202 Test: Non-Polar Ma Analysis Date: 06/20/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2243
pH Paper 0-14	W2186
Sodium Sulfate	EP1754
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP52823
LCSWD	10.00 ML	WP54439
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 71 Dessicator Time In : 11:00
1.0000 gram Balance: 1.0003 In Time: 10:15
Bal Check Time: 09:10

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 70 Dessicator Time Out: 12:45
1.0000 gram Balance: 1.0005 Out Time: 12:10
Bal Check Time: 13:25

LB88202

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POPLAR I3738

WorkList ID : 99782

Date : 6/20/2017 8:18:51 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3738-02	Non-Polar Material	Conc H2SO4 to	CORE01	M22	SEWER-OUTLET-GRAB	06/15/2017	1664A

Date/Time 06/20/17 8:18 AM
Received by: JP
Relinquished by: JP

Date/Time 06/20/17 10:00 AM
Received by: JP
Relinquished by: JP