



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88203

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/19/2017	13:30
2	CAL2	1	Water			7.00	06/19/2017	13:31
3	CAL3	1	Water			10.03	06/19/2017	13:33
4	ICV	1	Water			7.01	06/19/2017	13:35
5	CCV1	1	Water			2.01	06/19/2017	13:37
6	I3759-03	1	Solid	20.03	20	8.72	06/19/2017	13:38
7	I3759-03DUP	1	Solid	20.04	20	8.73	06/19/2017	13:40
8	I3762-03	1	Solid	20.01	20	7.36	06/19/2017	13:42
9	I3777-01	1	Solid	20.03	20	9.31	06/19/2017	13:43
10	CCV2	1	Water			12.01	06/19/2017	13:45

2B88203

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3777 WorkList ID : 99766 Date : 6/19/2017 1:01:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3777-01	Corrosivity	Cool 4 deg C	PARS15	M42	GHD-WC-S-1-0-52	06/16/2017	9045C
06/21/2017	Solid	I3759-03	Corrosivity	Cool 4 deg C	PSEG05	M51	SU-02-061617	06/16/2017	9045C
06/21/2017	Solid	I3762-03	Corrosivity	Cool 4 deg C	PSEG05	N61	SA-RO-0033-061617	06/16/2017	9045C

Date/Time 06/19/17 1:01 PM
Received by: JP
Relinquished by: JP

Date/Time 06/19/17 5:00 PM
Received by: JP
Relinquished by: JP