

LB 88204


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35500



A= 0.0000e+000

B= 2.7696e-004

C= 1.6513e-002

Rho= 0.9993826

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	-40	0.000	-40				
0.2	0.200	0.231	0.031	774	0.0 %	774				
2.5	2.500	2.319	-0.181	8312	0.0 %	8312				
5.0	5.000	5.145	0.145	18518	0.0 %	18518				
7.5	7.500	7.651	0.151	27567	0.0 %	27567				
10.0	10.000	9.848	-0.152	35500	0.0 %	35500				

LB88204 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 20 Jun 2017 09:13:00							
Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1	50	-40	-	0.0000	7470A	PPB	20 Jun 2017 09:29:38	S	Std		
0.2 - 1	50.2	774	-	0.2000	7470A	PPB	20 Jun 2017 09:31:46	S	Std		
2.5 - 1	52.5	8312	-	2.5000	7470A	PPB	20 Jun 2017 09:33:55	S	Std		
5.0 - 1	55	18518	-	5.0000	7470A	PPB	20 Jun 2017 09:36:03	S	Std		
7.5 - 1	57.5	27567	-	7.5000	7470A	PPB	20 Jun 2017 09:38:11	S	Std		
10.0 - 1	510	35500	-	10.0000	7470A	PPB	20 Jun 2017 09:45:10	S	Std		
ICV07 - 1	ICV07	14482	4.0274	-	7470A	PPB	20 Jun 2017 09:50:15	U	SMPL		
ICB07 - 1	ICB07	-148	-0.0245	-	7470A	PPB	20 Jun 2017 09:52:21	U	SMPL		
CCV86 - 1	CCV86	17654	4.9059	-	7470A	PPB	20 Jun 2017 09:54:28	U	SMPL		
CCB86 - 1	CCB86	-160	-0.0278	-	7470A	PPB	20 Jun 2017 09:56:35	U	SMPL		
CRA - 1	CRA	757	0.2262	-	7470A	PPB	20 Jun 2017 09:58:41	U	SMPL		
HighStd - 1	HighStd	35330	9.8014	-	7470A	PPB	20 Jun 2017 10:00:47	U	SMPL		
ChkStd - 1	ChkStd	25657	7.1224	-	7470A	PPB	20 Jun 2017 10:02:54	U	SMPL		
PB99922BL - 1	PBW	-108	-0.0134	-	7470A	PPB	20 Jun 2017 10:05:02	U	SMPL		
PB99922BS - 1	LCSW	14344	3.9892	-	7470A	PPB	20 Jun 2017 10:07:10	U	SMPL		
PB99922TB - 1	PB99879TB	-23	0.0101	-	7470A	PPB	20 Jun 2017 10:09:17	U	SMPL		
I3688-19 - 1	G23-0-1.5	-13	0.0129	-	7470A	PPB	20 Jun 2017 10:11:25	U	SMPL		
I3688-19DUP - 1	G23-0-1.5DUP	-14	0.0126	-	7470A	PPB	20 Jun 2017 10:13:33	U	SMPL		
I3688-19MS - 1	G23-0-1.5MS	13732	3.8197	-	7470A	PPB	20 Jun 2017 10:15:42	U	SMPL		
I3688-19MSD - 1	G23-0-1.5MSD	12687	3.5303	-	7470A	PPB	20 Jun 2017 10:17:49	U	SMPL		
CCV87 - 1	CCV87	17358	4.8239	-	7470A	PPB	20 Jun 2017 10:19:56	U	SMPL		
CCB87 - 1	CCB87	-125	-0.0181	-	7470A	PPB	20 Jun 2017 10:22:03	U	SMPL		
I3688-20 - 1	G23-1.5-3	68	0.0353	-	7470A	PPB	20 Jun 2017 10:24:09	U	SMPL		
I3688-21 - 1	G17-0-1.5	93	0.0423	-	7470A	PPB	20 Jun 2017 10:26:16	U	SMPL		
I3688-22 - 1	G17-1.5-3	103	0.0450	-	7470A	PPB	20 Jun 2017 10:28:23	U	SMPL		
I3688-23 - 1	G20-0-1.5	49	0.0301	-	7470A	PPB	20 Jun 2017 10:30:30	U	SMPL		
I3688-24 - 1	G20-1.5-3	33	0.0257	-	7470A	PPB	20 Jun 2017 10:32:37	U	SMPL		
I3688-25 - 1	G16-0-1.5	9	0.0190	-	7470A	PPB	20 Jun 2017 10:34:44	U	SMPL		
I3688-26 - 1	G16-1.5-3	5	0.0179	-	7470A	PPB	20 Jun 2017 10:36:53	U	SMPL		
I3688-27 - 1	G15-1.5-3	-10	0.0137	-	7470A	PPB	20 Jun 2017 10:39:01	U	SMPL		
I3688-28 - 1	G25-0-1.5	-34	0.0071	-	7470A	PPB	20 Jun 2017 10:41:10	U	SMPL		
I3688-29 - 1	G25-1.5-3	-11	0.0135	-	7470A	PPB	20 Jun 2017 10:43:17	U	SMPL		
CCV88 - 1	CCV88	17747	4.9317	-	7470A	PPB	20 Jun 2017 10:45:23	U	SMPL		
CCB88 - 1	CCB88	-21	0.0107	-	7470A	PPB	20 Jun 2017 10:47:30	U	SMPL		
I3688-30 - 1	G15-0-1.5	39	0.0273	-	7470A	PPB	20 Jun 2017 10:49:37	U	SMPL		
I3688-31 - 1	G13-0-1.5	-12	0.0132	-	7470A	PPB	20 Jun 2017 10:51:43	U	SMPL		
I3688-32 - 1	G13-1.5-3	-12	0.0132	-	7470A	PPB	20 Jun 2017 10:53:51	U	SMPL		
I3688-33 - 1	G12-0-1.5	-24	0.0099	-	7470A	PPB	20 Jun 2017 10:55:58	U	SMPL		
I3688-34 - 1	G12-1.5-3	-30	0.0082	-	7470A	PPB	20 Jun 2017 10:58:05	U	SMPL		
I3688-35 - 1	G11-0-1.5	-74	-0.0040	-	7470A	PPB	20 Jun 2017 11:00:13	U	SMPL		
I3688-36 - 1	G11-1.5-3	-61	-0.0004	-	7470A	PPB	20 Jun 2017 11:02:21	U	SMPL		
PB99923BL - 1	PBW	-25	0.0096	-	7470A	PPB	20 Jun 2017 11:04:29	U	SMPL		
PB99923BS - 1	LCSW	14502	4.0329	-	7470A	PPB	20 Jun 2017 11:06:38	U	SMPL		
PB99923TB - 1	PB99864TB	-23	0.0101	-	7470A	PPB	20 Jun 2017 11:08:44	U	SMPL		
CCV89 - 1	CCV89	17990	4.9990	-	7470A	PPB	20 Jun 2017 11:10:51	U	SMPL		
CCB89 - 1	CCB89	-58	0.0004	-	7470A	PPB	20 Jun 2017 11:12:58	U	SMPL		
I3722-02 - 1	C-5	-48	0.0032	-	7470A	PPB	20 Jun 2017 11:15:05	U	SMPL		
I3722-02DUP - 1	C-5DUP	-45	0.0040	-	7470A	PPB	20 Jun 2017 11:17:12	U	SMPL		
I3722-02MS - 1	C-5MS	13354	3.7150	-	7470A	PPB	20 Jun 2017 11:19:19	U	SMPL		
I3722-02MSD - 1	C-5MSD	13590	3.7804	-	7470A	PPB	20 Jun 2017 11:21:26	U	SMPL		
I3722-04 - 1	C-7	-20	0.0110	-	7470A	PPB	20 Jun 2017 11:23:34	U	SMPL		
I3722-06 - 1	C-8	-10	0.0137	-	7470A	PPB	20 Jun 2017 11:25:42	U	SMPL		
I3722-08 - 1	C-9	-11	0.0135	-	7470A	PPB	20 Jun 2017 11:27:50	U	SMPL		
I3722-10 - 1	C1-A	-59	0.0002	-	7470A	PPB	20 Jun 2017 11:29:58	U	SMPL		
I3722-12 - 1	C1-B	-61	-0.0004	-	7470A	PPB	20 Jun 2017 11:32:07	U	SMPL		
I3722-14 - 1	C4-A	-47	0.0035	-	7470A	PPB	20 Jun 2017 11:34:14	U	SMPL		
CCV90 - 1	CCV90	17663	4.9084	-	7470A	PPB	20 Jun 2017 11:36:21	U	SMPL		
CCB90 - 1	CCB90	-27	0.0090	-	7470A	PPB	20 Jun 2017 11:38:28	U	SMPL		
I3722-16 - 1	C4-B	-14	0.0126	-	7470A	PPB	20 Jun 2017 11:40:35	U	SMPL		
I3722-18 - 1	C6-A	-20	0.0110	-	7470A	PPB	20 Jun 2017 11:42:42	U	SMPL		
I3722-20 - 1	C6-B	-31	0.0079	-	7470A	PPB	20 Jun 2017 11:44:50	U	SMPL		
I3722-22 - 1	C2-A	-55	0.0013	-	7470A	PPB	20 Jun 2017 11:46:57	U	SMPL		
I3722-24 - 1	C2-B	-39	0.0057	-	7470A	PPB	20 Jun 2017 11:49:04	U	SMPL		
I3758-01 - 1	R1-TP-3-D34	-40	0.0054	-	7470A	PPB	20 Jun 2017 11:51:11	U	SMPL		
I3758-02 - 1	R1-TP5-D34	-37	0.0063	-	7470A	PPB	20 Jun 2017 11:53:19	U	SMPL		
I3758-03 - 1	R1-TP10-SS4	7	0.0185	-	7470A	PPB	20 Jun 2017 11:55:28	U	SMPL		
I3759-03 - 1	SU-02-061617	10	0.0193	-	7470A	PPB	20 Jun 2017 11:57:36	U	SMPL		
PB99924BL - 1	PBW	-4	0.0154	-	7470A	PPB	20 Jun 2017 11:59:43	U	SMPL		
CCV91 - 1	CCV91	17914	4.9779	-	7470A	PPB	20 Jun 2017 12:01:51	U	SMPL		
CCB91 - 1	CCB91	-83	-0.0065	-	7470A	PPB	20 Jun 2017 12:03:58	U	SMPL		
PB99924BS - 1	LCSW	13749	3.8244	-	7470A	PPB	20 Jun 2017 12:06:06	U	SMPL		
PB99924TB - 1	PB99880TB	-99	-0.0109	-	7470A	PPB	20 Jun 2017 12:08:13	U	SMPL		
I3736-13 - 1	G24-0-1.5	-36	0.0065	-	7470A	PPB	20 Jun 2017 12:10:21	U	SMPL		
I3736-13DUP - 1	G24-0-1.5DUP	4	0.0176	-	7470A	PPB	20 Jun 2017 12:12:28	U	SMPL		
I3736-13MS - 1	G24-0-1.5MS	13259	3.6887	-	7470A	PPB	20 Jun 2017 12:14:35	U	SMPL		
I3736-13MSD - 1	G24-0-1.5MSD	13632	3.7920	-	7470A	PPB	20 Jun 2017 12:16:43	U	SMPL		
I3736-14 - 1	G24-1.5-3	-39	0.0057	-	7470A	PPB	20 Jun 2017 12:18:51	U	SMPL		
I3736-15 - 1	G22-0-1.5	-39	0.0057	-	7470A	PPB	20 Jun 2017 12:20:59	U	SMPL		
I3736-16 - 1	G22-1.5-3	-39	0.0057	-	7470A	PPB	20 Jun 2017 12:23:08	U	SMPL		
I3736-17 - 1	G19-0-1.5	-22	0.0104	-	7470A	PPB	20 Jun 2017 12:25:15	U	SMPL		
CCV92 - 1	CCV92	18333	5.0940	-	7470A	PPB	20 Jun 2017 12:27:23	U	SMPL		

LB88204 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Jun 2017 09:13:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB92 - 1	CCB92	-45	0.0040		-7470A	PPB	20 Jun 2017 12:29:31	U	SMPL
I3736-18 - 1	G19-1.5-3	20	0.0221		-7470A	PPB	20 Jun 2017 12:31:38	U	SMPL
I3736-19 - 1	G14-0-1.5	-10	0.0137		-7470A	PPB	20 Jun 2017 12:33:46	U	SMPL
I3736-20 - 1	G14-1.5-3	-38	0.0060		-7470A	PPB	20 Jun 2017 12:35:54	U	SMPL
I3736-21 - 1	G18-0-1.5	-30	0.0082		-7470A	PPB	20 Jun 2017 12:38:02	U	SMPL
I3736-22 - 1	G18-1.5-3	-43	0.0046		-7470A	PPB	20 Jun 2017 12:40:09	U	SMPL
I3736-23 - 1	G21-0-1.5	-39	0.0057		-7470A	PPB	20 Jun 2017 12:42:17	U	SMPL
I3736-24 - 1	G21-1.5-3	-30	0.0082		-7470A	PPB	20 Jun 2017 12:44:25	U	SMPL
PB99925BL - 1	PBW	0	0.0165		-7470A	PPB	20 Jun 2017 12:46:33	U	SMPL
PB99925BS - 1	LCSW	13045	3.6294		-7470A	PPB	20 Jun 2017 12:48:42	U	SMPL
PB99925TB - 1	PB99884TB	-61	-0.0004		-7470A	PPB	20 Jun 2017 12:50:50	U	SMPL
CCV93 - 1	CCV93	18540	5.1513		-7470A	PPB	20 Jun 2017 12:52:58	U	SMPL
CCB93 - 1	CCB93	-95	-0.0098		-7470A	PPB	20 Jun 2017 12:55:06	U	SMPL
I3749-03 - 1	TILCON-MH-001	-96	-0.0101		-7470A	PPB	20 Jun 2017 12:57:13	U	SMPL
I3749-03DUP - 1	TILCON-MH-001DUP	-84	-0.0068		-7470A	PPB	20 Jun 2017 12:59:21	U	SMPL
I3749-03MS - 1	TILCON-MH-001MS	13506	3.7571		-7470A	PPB	20 Jun 2017 13:01:29	U	SMPL
I3749-03MSD - 1	TILCON-MH-001MSD	13566	3.7737		-7470A	PPB	20 Jun 2017 13:03:37	U	SMPL
I3749-04 - 1	TILCON-PL-001	-136	-0.0212		-7470A	PPB	20 Jun 2017 13:05:44	U	SMPL
CCV94 - 1	CCV94	18386	5.1087		-7470A	PPB	20 Jun 2017 13:07:52	U	SMPL
CCB94 - 1	CCB94	-35	0.0068		-7470A	PPB	20 Jun 2017 13:10:00	U	SMPL
I3688-19LX5 - 1		28	0.0243		-7470A	PPB	20 Jun 2017 13:12:08	U	SMPL
I3688-19A - 1		13883	3.8615		-7470A	PPB	20 Jun 2017 13:14:16	U	SMPL
I3722-02LX5 - 1		-49	0.0029		-7470A	PPB	20 Jun 2017 13:16:24	U	SMPL
I3722-02A - 1		14074	3.9144		-7470A	PPB	20 Jun 2017 13:18:32	U	SMPL
I3736-13LX5 - 1		-49	0.0029		-7470A	PPB	20 Jun 2017 13:20:40	U	SMPL
I3736-13A - 1		13971	3.8859		-7470A	PPB	20 Jun 2017 13:22:47	U	SMPL
I3749-03LX5 - 1		-51	0.0024		-7470A	PPB	20 Jun 2017 13:24:55	U	SMPL
I3749-03A - 1		13897	3.8654		-7470A	PPB	20 Jun 2017 13:27:03	U	SMPL
CCV95 - 1	CCV95	17339	4.8187		-7470A	PPB	20 Jun 2017 13:29:11	U	SMPL
CCB95 - 1	CCB95	-79	-0.0054		-7470A	PPB	20 Jun 2017 13:31:19	U	SMPL