


Chemtech Consulting Group

Analytical Review Report

Date Printed : 6/22/17
 Analyst : AK
 Data File : LB88206.csv

Approved By : 
 Approved Date : 6/22/17
 Worksheet # :


m9056-A-Inorganic Anions-13

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1	
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Chloride												
STD1	STD1				W	6/12/17						
Chloride		PASS	0.000		0.000						mg/L	
STD2	STD2				W	6/12/17						
Chloride		PASS	0.202		0.202						mg/L	
STD3	STD3				W	6/12/17						
Chloride		PASS	0.612		0.612						mg/L	
STD4	STD4				W	6/12/17						
Chloride		PASS	1.149		1.15						mg/L	
STD5	STD5				W	6/12/17						
Chloride		PASS	2.994		2.99						mg/L	
STD6	STD6				W	6/12/17						
Chloride		PASS	5.930		5.93						mg/L	
STD7	STD7				W	6/12/17						
Chloride		PASS	7.563		7.56						mg/L	
ICV1	ICV1				W	6/12/17						
Chloride		PASS	2.949		2.95	98.0	90	110			mg/L	
ICB1	ICB1				W	6/12/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
CCV1	CCV1				W	6/19/17						
Chloride		PASS	2.934		2.93	98.0	90	110			mg/L	
CCB1	CCB1				W	6/19/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
LB88206BLS	LB88206BLS				S	6/19/17						
Chloride		PASS	0.000		0.000			+/-0.0320			mg/Kg	
LB88206BSS	LB88206BSS				S	6/19/17						
Chloride		PASS	3.005		59.90	100.0	90.00	110.00			mg/Kg	
I3774-01	CHLOR-1			1	S	6/19/17						
Chloride		OR	63.614		1250						mg/Kg	
I3774-02	CHLOR-2			1	S	6/19/17						
Chloride		OR	371.092		6830						mg/Kg	
I3774-03	CHLOR-3			1	S	6/19/17						
Chloride		OR	9.658		1180						mg/Kg	
I3774-01D	CHLOR-1D			1	S	6/19/17						
Chloride		PASS	63.978		1260.00				0.8	20	mg/Kg	
I3774-01S	CHLOR-1S			1	S	6/19/17						
Chloride		FAIL	61.324		1210.0	-68.0	80	120			mg/Kg	
I3774-01SD	CHLOR-1SD			1	S	6/19/17						
Chloride		FAIL	61.105		1210.0	-67.0	80	120	0	15	mg/Kg	
CCV2	CCV2				W	6/19/17						
Chloride		PASS	2.999		3.00	100.0	90	110			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 6/22/17
Analyst : AK
Data File : LB88206.csv

Approved By : 
Approved Date :
Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			
Parameter			PPB		Final Conc	%Rec	Method	Method			
							LCL	UCL	RPD	Max RPD	Units
Chloride											
CCB2	CCB2				W	6/19/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
I3774-01DL	CHLOR-1DL			10	S	6/19/17					
Chloride		PASS	6.331		1250						mg/Kg
I3774-02DL	CHLOR-2DL			50	S	6/20/17					
Chloride		PASS	5.673		5220						mg/Kg
I3774-03DL	CHLOR-3DL			2	S	6/20/17					
Chloride		PASS	4.701		1150						mg/Kg
CCV3	CCV3				W	6/20/17					
Chloride		PASS	2.981		2.98	99.0	90	110			mg/L
CCB3	CCB3				W	6/20/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L