

Analysis Method: SM4500-CL C

ANALYST: AMANDEEP

Parameter: Chloride

SUPERVISOR REVIEW BY: apatel

Run Number: LB88208

Constant: 35450

Normality1: 0.0141

Reagent/Standard	Lot/Log #
Chloride LCS std, 100PPM	WP55779
Indicator-acidifier reagent - Chloride	WP55324
MERCURIC NITRATE, 0.0141 N 4L	W2051

Seq	Lab ID	True Value (mg/L)	Dilution	Sample Volume (mL)	Titrant Initial (ml)	Titrant Final (ml)	Titrant Diff. (ml)	Result (mg/L)	Anal Date	Anal Time
1	LB88208BL		1	50	0.00	0.10	0.10	0.00	06/20/2017	09:30
2	LB88208BS	100	1	10	0.10	2.18	2.08	98.97	06/20/2017	09:41
3	I3760-01		1	10	0.98	2.86	1.88	88.97	06/20/2017	09:56
4	I3760-01DUP		1	10	2.86	4.80	1.94	91.97	06/20/2017	10:06
5	I3760-01MS	100	1	10	4.80	8.50	3.70	179.94	06/20/2017	10:15
6	I3760-01MSD	100	1	10	0.00	3.76	3.76	182.94	06/20/2017	10:24

Titrant Diff = TitrantFinal - TitrantInitial

Result = (Titrant Diff - Row Blank Value * Normality)) * Constant /

W088208

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CL061917

WorkList ID : 99758

Date : 6/19/2017 11:16:25 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3760-01	Chloride	Cool 4 deg C	HDR103	N22	TZBUD-20170616	06/16/2017	SM4500-CLC

Date/Time 6-20-17 9:18
 Received by: J.S
 Relinquished by: OK

Date/Time 6-20-17 11:00
 Received by: AK
 Relinquished by: J.S