

7470A


LB88214

INSTRUMENT ID: CV4

Linear

 μ Abs.:

35826



A= 0.0000e+000

B= 2.7978e-004

C= 1.6764e-002

Rho= 0.9999474

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	-28	0.000	-28				
0.2	0.200	0.195	-0.005	636	0.0 %	636				
2.5	2.500	2.547	0.047	9044	0.0 %	9044				
5.0	5.000	4.938	-0.062	17590	0.0 %	17590				
7.5	7.500	7.471	-0.029	26644	0.0 %	26644				
10.0	10.000	10.040	0.040	35826	0.0 %	35826				

LB88214 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 20 Jun 2017 14:09:53				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-28	-	0.00007470A	PPB	20 Jun 2017 14:11:39	S	Std
0.2 - 1	50.2	636	-	0.20007470A	PPB	20 Jun 2017 14:13:47	S	Std
2.5 - 1	52.5	9044	-	2.50007470A	PPB	20 Jun 2017 14:15:56	S	Std
5.0 - 1	55	17590	-	5.00007470A	PPB	20 Jun 2017 14:18:04	S	Std
7.5 - 1	57.5	26644	-	7.50007470A	PPB	20 Jun 2017 14:20:12	S	Std
10.0 - 1	510	35826	-	10.00007470A	PPB	20 Jun 2017 14:22:19	S	Std
ICV08 - 1	ICV08	14425	4.0526	-7470A	PPB	20 Jun 2017 14:27:46	U	SMPL
ICB08 - 1	ICB08	-69	-0.0025	-7470A	PPB	20 Jun 2017 14:29:52	U	SMPL
CCV96 - 1	CCV96	17848	5.0102	-7470A	PPB	20 Jun 2017 14:31:59	U	SMPL
CCB96 - 1	CCB96	-82	-0.0062	-7470A	PPB	20 Jun 2017 14:34:05	U	SMPL
CRA - 1	CRA	624	0.1913	-7470A	PPB	20 Jun 2017 14:36:11	U	SMPL
HighStd - 1	HighStd	36031	10.0974	-7470A	PPB	20 Jun 2017 14:38:18	U	SMPL
ChkStd - 1	ChkStd	24752	6.9418	-7470A	PPB	20 Jun 2017 14:40:25	U	SMPL
PB99942BL - 1	PBW	-117	-0.0160	-7470A	PPB	20 Jun 2017 14:42:33	U	SMPL
PB99942BS - 1	LCSW	13306	3.7395	-7470A	PPB	20 Jun 2017 14:44:40	U	SMPL
PB99942TB - 1	PB99926TB	-92	-0.0090	-7470A	PPB	20 Jun 2017 14:46:48	U	SMPL
I3634-05 - 1	SP-KEARNY-COMP	-51	0.0025	-7470A	PPB	20 Jun 2017 14:48:56	U	SMPL
I3634-05DUP - 1	SP-KEARNY-COMPDUP	-26	0.0095	-7470A	PPB	20 Jun 2017 14:51:04	U	SMPL
I3634-05MS - 1	SP-KEARNY-COMPMS	13994	3.9320	-7470A	PPB	20 Jun 2017 14:53:13	U	SMPL
I3634-05MSD - 1	SP-KEARNY-COMPMSD	13738	3.8603	-7470A	PPB	20 Jun 2017 14:55:20	U	SMPL
CCV97 - 1	CCV97	17442	4.8966	-7470A	PPB	20 Jun 2017 14:57:27	U	SMPL
CCB97 - 1	CCB97	-100	-0.0112	-7470A	PPB	20 Jun 2017 14:59:33	U	SMPL
I3634-05LX5 - 1		-55	0.0014	-7470A	PPB	20 Jun 2017 15:01:40	U	SMPL
I3634-05A - 1		14279	4.0117	-7470A	PPB	20 Jun 2017 15:03:47	U	SMPL
CCV98 - 1	CCV98	17627	4.9484	-7470A	PPB	20 Jun 2017 15:05:54	U	SMPL
CCB98 - 1	CCB98	-86	-0.0073	-7470A	PPB	20 Jun 2017 15:08:01	U	SMPL