



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88224

Temp.: 21.6 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/20/2017	15:00
2	CAL2	1	Water			7.01	06/20/2017	15:01
3	CAL3	1	Water			10.02	06/20/2017	15:03
4	ICV	1	Water			7.00	06/20/2017	15:05
5	CCV1	1	Water			2.01	06/20/2017	15:07
6	I3800-01	1	Solid	20.02	20	9.42	06/20/2017	15:08
7	I3800-01DUP	1	Solid	20.03	20	9.43	06/20/2017	15:10
8	CCV2	1	Water			12.01	06/20/2017	15:12

2B88224

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3800 WorkList ID : 99828 Date : 6/20/2017 2:39:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/20/2017	Solid	I3800-01-X	pH	Cool 4 deg C	GLOB02	N31	CRUSHED-STONE-DUST	06/20/2017	9045C

Date/Time	06/20/17 2:40 PM	06/20/17	4:30 PM
Received by:	JS	JS	
Relinquished by:	JS	JS	