


7471A

LB88227
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43376



A= 0.0000e+000

B= 2.2740e-004

C= 3.3579e-002

Rho= 0.9997048

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	-19	0.000	-19				
0.2	0.200	0.200	-0.000	731	0.0 %	731				
2.5	2.500	2.431	-0.069	10544	0.0 %	10544				
5.0	5.000	4.967	-0.033	21694	0.0 %	21694				
7.5	7.500	7.676	0.176	33608	0.0 %	33608				
10.0	10.000	9.897	-0.103	43376	0.0 %	43376				

LB88227 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 20 Jun 2017 16:35:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0000	7471A	PPB	20 Jun 2017 16:44:48	S	Std
0.2 - 1	50.2	731	-	0.2000	7471A	PPB	20 Jun 2017 16:47:06	S	Std
2.5 - 1	52.5	10544	-	2.5000	7471A	PPB	20 Jun 2017 16:49:23	S	Std
5.0 - 1	55	21694	-	5.0000	7471A	PPB	20 Jun 2017 16:51:40	S	Std
7.5 - 1	57.5	33608	-	7.5000	7471A	PPB	20 Jun 2017 16:53:58	S	Std
10.0 - 1	510	43376	-	10.0000	7471A	PPB	20 Jun 2017 16:56:14	S	Std
ICV09 - 1	ICV09	16778	3.8488	-	7471A	PPB	20 Jun 2017 16:58:53	U	SMPL
ICB09 - 1	ICB09	-86	0.0140	-	7471A	PPB	20 Jun 2017 17:01:09	U	SMPL
CCV01 - 1	CCV01	21560	4.9362	-	7471A	PPB	20 Jun 2017 17:03:24	U	SMPL
CCB01 - 1	CCB01	-104	0.0099	-	7471A	PPB	20 Jun 2017 17:05:40	U	SMPL
CRA - 1	CRA	710	0.1950	-	7471A	PPB	20 Jun 2017 17:07:56	U	SMPL
HighStd - 1	HighStd	42844	9.7761	-	7471A	PPB	20 Jun 2017 17:10:12	U	SMPL
ChkStd - 1	ChkStd	29807	6.8116	-	7471A	PPB	20 Jun 2017 17:12:28	U	SMPL
PB99927BL - 1	PBS	-135	0.0029	-	7471A	PPB	20 Jun 2017 17:14:44	U	SMPL
PB99927BS - 1	LCSS	16768	3.8466	-	7471A	PPB	20 Jun 2017 17:17:01	U	SMPL
I3762-01 - 1	SA-RO-0033-061617	4700	1.1023	-	7471A	PPB	20 Jun 2017 17:19:18	U	SMPL
I3775-02 - 1	GHD-SB-020304-0-11	7022	1.6304	-	7471A	PPB	20 Jun 2017 17:21:35	U	SMPL
I3775-02DUP - 1	GHD-SB-020304-0-11DUP	6056	1.4107	-	7471A	PPB	20 Jun 2017 17:23:53	U	SMPL
I3775-03 - 1	I3775-02MS	23769	5.4385	-	7471A	PPB	20 Jun 2017 17:26:11	U	SMPL
I3775-04 - 1	I3775-02MSD	23681	5.4185	-	7471A	PPB	20 Jun 2017 17:28:27	U	SMPL
CCV02 - 1	CCV02	21891	5.0115	-	7471A	PPB	20 Jun 2017 17:30:43	U	SMPL
CCB02 - 1	CCB02	-76	0.0163	-	7471A	PPB	20 Jun 2017 17:32:59	U	SMPL
I3775-05 - 1	GHD-SB-1020304-0-11	9095	2.1017	-	7471A	PPB	20 Jun 2017 17:35:15	U	SMPL
I3775-08 - 1	GHD-SB-01-0-52	8648	2.0001	-	7471A	PPB	20 Jun 2017 17:37:31	U	SMPL
I3775-02LX5 - 1		1313	0.3321	-	7471A	PPB	20 Jun 2017 17:39:46	U	SMPL
I3775-02A - 1		24372	5.5757	-	7471A	PPB	20 Jun 2017 17:42:03	U	SMPL
CCV03 - 1	CCV03	22014	5.0395	-	7471A	PPB	20 Jun 2017 17:44:19	U	SMPL
CCB03 - 1	CCB03	-76	0.0163	-	7471A	PPB	20 Jun 2017 17:46:36	U	SMPL