



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88233

Temp.: 21.6 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/20/2017	15:00
2	CAL2	1	Water			7.01	06/20/2017	15:01
3	CAL3	1	Water			10.02	06/20/2017	15:03
4	ICV	1	Water			7.00	06/20/2017	15:05
5	CCV1	1	Water			2.01	06/20/2017	15:30
6	I3781-07	1	Solid	20.02	20	8.90	06/20/2017	15:31
7	I3781-08	1	Solid	20.04	20	8.00	06/20/2017	15:33
8	I3792-03	1	Solid	20.01	20	8.46	06/20/2017	15:34
9	I3799-16	1	Solid	20.02	20	8.16	06/20/2017	15:36
10	I3799-17	1	Solid	20.04	20	7.24	06/20/2017	15:39
11	I3799-18	1	Solid	20.02	20	6.67	06/20/2017	15:40
12	I3799-19	1	Solid	20.05	20	7.62	06/20/2017	15:42
13	I3799-20	1	Solid	20.03	20	7.80	06/20/2017	15:43
14	I3799-21	1	Solid	20.02	20	7.53	06/20/2017	15:45
15	I3799-22	1	Solid	20.03	20	8.30	06/20/2017	15:46
16	CCV2	1	Water			12.01	06/20/2017	15:48
17	I3802-03	1	Solid	20.01	20	7.53	06/20/2017	15:50
18	I3802-03DUP	1	Solid	20.03	20	7.54	06/20/2017	15:51
19	CCV3	1	Water			12.01	06/20/2017	15:52

1788233

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3792 WorkList ID : 99826 Date : 6/20/2017 1:50:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3781-07	Corrosivity	Cool 4 deg C	LIRO01	M42	SB-09-COMP	06/19/2017	9045C
	Solid	I3781-08	Corrosivity	Cool 4 deg C	LIRO01	M42	SB-01-COMP	06/19/2017	9045C
06/22/2017	Solid	I3792-03	Corrosivity	Cool 4 deg C	PSEG05	N21	OR-03-061917	06/19/2017	9045C
	Solid	I3799-16	Corrosivity	Cool 4 deg C	LIRO01	N41	SB-02-COMP	06/19/2017	9045C
	Solid	I3799-17	Corrosivity	Cool 4 deg C	LIRO01	N41	SB-03-COMP	06/19/2017	9045C
	Solid	I3799-18	Corrosivity	Cool 4 deg C	LIRO01	N41	SB-04-COMP	06/19/2017	9045C
	Solid	I3799-19	Corrosivity	Cool 4 deg C	LIRO01	N41	SB-05-COMP	06/20/2017	9045C
	Solid	I3799-20	Corrosivity	Cool 4 deg C	LIRO01	N41	SB-06-COMP	06/20/2017	9045C
	Solid	I3799-21	Corrosivity	Cool 4 deg C	LIRO01	N41	SB-07-COMP	06/20/2017	9045C
	Solid	I3799-22	Corrosivity	Cool 4 deg C	LIRO01	N41	SB-08-COMP	06/20/2017	9045C
06/23/2017	Solid	I3802-03	Corrosivity	Cool 4 deg C	PSEG05	N33	OK-02-062017	06/20/2017	9045C

Date/Time 06/20/17 1:50 PM
 Received by: J.C.
 Relinquished by: J.C.

Date/Time 06/20/17 6:00 PM
 Received by: J.C.
 Relinquished by: J.C.