


7470A

LB 88237
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35768



A= 0.0000e+000

B= 2.7799e-004

C= 3.1898e-002

Rho= 0.9999741

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	-56	0.000	-56				
0.2	0.200	0.195	-0.005	586	0.0 %	586				
2.5	2.500	2.480	-0.020	8807	0.0 %	8807				
5.0	5.000	4.983	-0.017	17809	0.0 %	17809				
7.5	7.500	7.551	0.051	27049	0.0 %	27049				
10.0	10.000	9.975	-0.025	35768	0.0 %	35768				

LB88237 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Jun 2017 08:34:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-56	-	0.0000	7470A	PPB	21 Jun 2017 08:42:55	S	Std
0.2 - 1	30.2	586	-	0.2000	7470A	PPB	21 Jun 2017 08:45:03	S	Std
2.5 - 1	32.5	8807	-	2.5000	7470A	PPB	21 Jun 2017 08:47:12	S	Std
5.0 - 1	55	17809	-	5.0000	7470A	PPB	21 Jun 2017 08:49:20	S	Std
7.5 - 1	57.5	27049	-	7.5000	7470A	PPB	21 Jun 2017 08:51:28	S	Std
10.0 - 1	510	35768	-	10.0000	7470A	PPB	21 Jun 2017 08:53:35	S	Std
ICV10 - 1	ICV10	14170	3.9710	-	7470A	PPB	21 Jun 2017 08:56:02	U	SMPL
ICB10 - 1	ICB10	-101	0.0038	-	7470A	PPB	21 Jun 2017 08:58:08	U	SMPL
CCV04 - 1	CCV04	18731	5.2389	-	7470A	PPB	21 Jun 2017 09:00:14	U	SMPL
CCB04 - 1	CCB04	-140	-0.0070	-	7470A	PPB	21 Jun 2017 09:02:21	U	SMPL
CRA - 1	CRA	599	0.1984	-	7470A	PPB	21 Jun 2017 09:04:28	U	SMPL
HighStd - 1	HighStd	36166	10.0856	-	7470A	PPB	21 Jun 2017 09:06:34	U	SMPL
ChkStd - 1	ChkStd	24448	6.8281	-	7470A	PPB	21 Jun 2017 09:08:41	U	SMPL
PB99959BL - 1	PBW	-126	-0.0031	-	7470A	PPB	21 Jun 2017 09:10:49	U	SMPL
PB99959BS - 1	LCSW	13879	3.8901	-	7470A	PPB	21 Jun 2017 09:12:56	U	SMPL
PB99959TB - 1	PB99919TB	-99	0.0044	-	7470A	PPB	21 Jun 2017 09:15:04	U	SMPL
I3762-03 - 1	SA-RO-0033-061617	-80	0.0097	-	7470A	PPB	21 Jun 2017 09:17:12	U	SMPL
I3762-03DUP - 1	SA-RO-0033-061617DUP	-63	0.0144	-	7470A	PPB	21 Jun 2017 09:19:21	U	SMPL
I3762-03MS - 1	SA-RO-0033-061617MS	13001	3.6460	-	7470A	PPB	21 Jun 2017 09:21:30	U	SMPL
I3762-03MSD - 1	SA-RO-0033-061617MSD	13015	3.6499	-	7470A	PPB	21 Jun 2017 09:23:36	U	SMPL
CCV05 - 1	CCV05	18395	5.1455	-	7470A	PPB	21 Jun 2017 09:25:43	U	SMPL
CCB05 - 1	CCB05	-102	0.0035	-	7470A	PPB	21 Jun 2017 09:27:50	U	SMPL
I3777-01 - 1	GHD-WC-S-1-0-52	39	0.0427	-	7470A	PPB	21 Jun 2017 09:29:57	U	SMPL
I3781-07 - 1	SB-09-COMP	-48	0.0186	-	7470A	PPB	21 Jun 2017 09:32:04	U	SMPL
I3781-08 - 1	SB-01-COMP	-66	0.0136	-	7470A	PPB	21 Jun 2017 09:34:10	U	SMPL
I3784-07 - 1	P.56-A-COMP	-52	0.0174	-	7470A	PPB	21 Jun 2017 09:36:17	U	SMPL
I3784-08 - 1	P.56-B-COMP	-60	0.0152	-	7470A	PPB	21 Jun 2017 09:38:25	U	SMPL
I3784-09 - 1	P.56-C-COMP	-67	0.0133	-	7470A	PPB	21 Jun 2017 09:40:33	U	SMPL
PB99964BL - 1	PBW	-63	0.0144	-	7470A	PPB	21 Jun 2017 09:42:41	U	SMPL
PB99964BS - 1	LCSW	13275	3.7222	-	7470A	PPB	21 Jun 2017 09:44:49	U	SMPL
I3775-06 - 1	GHD-FB061517	-146	-0.0087	-	7470A	PPB	21 Jun 2017 09:46:58	U	SMPL
I3776-01 - 1	GHD-GW-01	83488	23.2405	-	7470A	PPB	21 Jun 2017 09:49:05	U	SMPL
CCV06 - 1	CCV06	18330	5.1274	-	7470A	PPB	21 Jun 2017 09:51:12	U	SMPL
CCB06 - 1	CCB06	-109	0.0016	-	7470A	PPB	21 Jun 2017 09:53:20	U	SMPL
I3793-01 - 1	GW-1	5996	1.6987	-	7470A	PPB	21 Jun 2017 09:55:27	U	SMPL
I3793-01DUP - 1	GW-1DUP	6042	1.7115	-	7470A	PPB	21 Jun 2017 09:57:34	U	SMPL
I3793-01MS - 1	GW-1MS	20595	5.7571	-	7470A	PPB	21 Jun 2017 09:59:41	U	SMPL
I3793-01MSD - 1	GW-1MSD	20181	5.6420	-	7470A	PPB	21 Jun 2017 10:01:48	U	SMPL
I3793-02 - 1	GW-1	-144	-0.0081	-	7470A	PPB	21 Jun 2017 10:03:55	U	SMPL
I3762-03LX5 - 1		-36	0.0219	-	7470A	PPB	21 Jun 2017 10:06:03	U	SMPL
I3762-03A - 1		14291	4.0046	-	7470A	PPB	21 Jun 2017 10:08:11	U	SMPL
I3793-01LX5 - 1		1037	0.3202	-	7470A	PPB	21 Jun 2017 10:10:19	U	SMPL
I3793-01A - 1		20409	5.7053	-	7470A	PPB	21 Jun 2017 10:12:27	U	SMPL
I3776-01DLX5 - 1		15100	4.2295	-	7470A	PPB	21 Jun 2017 10:14:34	U	SMPL
CCV07 - 1	CCV07	18534	5.1841	-	7470A	PPB	21 Jun 2017 10:16:41	U	SMPL
CCB07 - 1	CCB07	-108	0.0019	-	7470A	PPB	21 Jun 2017 10:18:49	U	SMPL