


ISM02.4

LB 88241
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40577



A= 0.0000e+000

B= 2.4811e-004

C= -2.5559e-002

Rho= 0.9999694

Accept = Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.00	0.000	-0.005	-0.005	81	0.000	81					16
0.2	0.200	0.232	0.032	1040	0.0 %	1040					0
2.5	2.500	2.494	-0.006	10157	0.0 %	10157					0
5.0	5.000	4.976	-0.024	20161	0.0 %	20161					0
7.5	7.500	7.460	-0.040	30172	0.0 %	30172					1
10.0	10.000	10.042	0.042	40577	0.0 %	40577					0

NB
6/21/17

LB88241
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	81	-		0 ISM02.4	PPB	6/21/2017 10:42	Std
	0.2 S0.2	1040	-		0.2 ISM02.4	PPB	6/21/2017 10:44	Std
	2.5 S2.5	10157	-		2.5 ISM02.4	PPB	6/21/2017 10:47	Std
	5 S5	20161	-		5 ISM02.4	PPB	6/21/2017 10:49	Std
	7.5 S7.5	30172	-		7.5 ISM02.4	PPB	6/21/2017 10:51	Std
	10 S10	40577	-		10 ISM02.4	PPB	6/21/2017 10:53	Std
ICV	ICV	15933	3.9275 -		ISM02.4	PPB	6/21/2017 10:56	SMPL
ICB	ICB	47	-0.0139 -		ISM02.4	PPB	6/21/2017 10:59	SMPL
CCV008	CCV008	20203	4.9869 -		ISM02.4	PPB	6/21/2017 11:01	SMPL
CCB008	CCB008	16	-0.0216 -		ISM02.4	PPB	6/21/2017 11:03	SMPL
PB99965BL	PBW003	-117	-0.0546 -		ISM02.4	PPB	6/21/2017 11:05	SMPL
I3659-01	MBDC80	158	0.0136 -		ISM02.4	PPB	6/21/2017 11:08	SMPL
I3659-02	MBDC81	-111	-0.0531 -		ISM02.4	PPB	6/21/2017 11:10	SMPL
I3659-03	MBDC83	-44	-0.0365 -		ISM02.4	PPB	6/21/2017 11:12	SMPL
I3659-04	MBDC83D	-72	-0.0434 -		ISM02.4	PPB	6/21/2017 11:14	SMPL
I3659-05	MBDC83S	4124	0.9976 -		ISM02.4	PPB	6/21/2017 11:17	SMPL
I3659-06	MBDC75	-129	-0.0576 -		ISM02.4	PPB	6/21/2017 11:19	SMPL
I3659-07	MBDC76	-115	-0.0541 -		ISM02.4	PPB	6/21/2017 11:21	SMPL
I3659-08	MBDC84	-98	-0.0499 -		ISM02.4	PPB	6/21/2017 11:24	SMPL
I3659-09	MBDC57	-115	-0.0541 -		ISM02.4	PPB	6/21/2017 11:26	SMPL
I3659-10	MBDC67	-61	-0.0407 -		ISM02.4	PPB	6/21/2017 11:28	SMPL
I3659-11	MBDC70	-99	-0.0501 -		ISM02.4	PPB	6/21/2017 11:30	SMPL
I3659-12	MBDC74	-105	-0.0516 -		ISM02.4	PPB	6/21/2017 11:33	SMPL
I3659-13	MBDC63	-130	-0.0578 -		ISM02.4	PPB	6/21/2017 11:35	SMPL
I3659-14	MBDC64	-119	-0.0551 -		ISM02.4	PPB	6/21/2017 11:37	SMPL
I3659-15	MBDC65	-107	-0.0521 -		ISM02.4	PPB	6/21/2017 11:40	SMPL
I3659-16	MBDC71	-92	-0.0484 -		ISM02.4	PPB	6/21/2017 11:42	SMPL
I3659-17	MBDC77	-69	-0.0427 -		ISM02.4	PPB	6/21/2017 11:44	SMPL
I3659-18	MBDCC0	-120	-0.0553 -		ISM02.4	PPB	6/21/2017 11:46	SMPL
PB99966BL	PBW003	-65	-0.0417 -		ISM02.4	PPB	6/21/2017 11:49	SMPL
CCV009	CCV009	19948	4.9236 -		ISM02.4	PPB	6/21/2017 11:51	SMPL
CCB009	CCB009	22	-0.0201 -		ISM02.4	PPB	6/21/2017 11:53	SMPL
I3770-01	MA4X31	-101	-0.0506 -		ISM02.4	PPB	6/21/2017 11:56	SMPL
I3770-02	MA4X32	-94	-0.0489 -		ISM02.4	PPB	6/21/2017 11:58	SMPL
I3770-03	MA4X33	-84	-0.0464 -		ISM02.4	PPB	6/21/2017 12:00	SMPL
I3770-04	MA4X33D	-73	-0.0437 -		ISM02.4	PPB	6/21/2017 12:02	SMPL
I3770-05	MA4X33S	4264	1.0324 -		ISM02.4	PPB	6/21/2017 12:05	SMPL
I3770-06	MA4X34	-105	-0.0516 -		ISM02.4	PPB	6/21/2017 12:07	SMPL
I3770-07	MA4X35	-129	-0.0576 -		ISM02.4	PPB	6/21/2017 12:09	SMPL
I3770-08	MA4X36	-100	-0.0504 -		ISM02.4	PPB	6/21/2017 12:11	SMPL
I3770-09	MA4X43	-108	-0.0524 -		ISM02.4	PPB	6/21/2017 12:14	SMPL
I3770-10	MA4X44	-115	-0.0541 -		ISM02.4	PPB	6/21/2017 12:16	SMPL
I3770-11	MA4X47	-43	-0.0362 -		ISM02.4	PPB	6/21/2017 12:18	SMPL
I3770-12	MA4X48	-66	-0.0419 -		ISM02.4	PPB	6/21/2017 12:21	SMPL
I3770-13	MA4X49	-87	-0.0471 -		ISM02.4	PPB	6/21/2017 12:23	SMPL
I3770-14	MA4X56	-127	-0.0571 -		ISM02.4	PPB	6/21/2017 12:25	SMPL

LB88241
INSTRUMENT ID : CV1

I3770-15	MA4X57	-130	-0.0578 -	ISM02.4	PPB	6/21/2017 12:27	SMPL
I3770-16	MA4X58	-101	-0.0506 -	ISM02.4	PPB	6/21/2017 12:30	SMPL
I3770-17	MA4X59	-118	-0.0548 -	ISM02.4	PPB	6/21/2017 12:32	SMPL
I3770-18	MA4X67	-14	-0.029 -	ISM02.4	PPB	6/21/2017 12:34	SMPL
I3770-19	MA4X68	-68	-0.0424 -	ISM02.4	PPB	6/21/2017 12:37	SMPL
I3770-20	MA4X69	-71	-0.0432 -	ISM02.4	PPB	6/21/2017 12:39	SMPL
CCV010	CCV010	19960	4.9266 -	ISM02.4	PPB	6/21/2017 12:41	SMPL
CCB010	CCB010	12	-0.0226 -	ISM02.4	PPB	6/21/2017 12:43	SMPL
I3770-21	MA4X70	-102	-0.0509 -	ISM02.4	PPB	6/21/2017 12:46	SMPL
I3770-22	MA4X71	-152	-0.0633 -	ISM02.4	PPB	6/21/2017 12:48	SMPL
PB99967BL	PBW003	-78	-0.0449 -	ISM02.4	PPB	6/21/2017 12:50	SMPL
I3772-01	MA4X37	-78	-0.0449 -	ISM02.4	PPB	6/21/2017 12:53	SMPL
I3772-02	MA4X38	-66	-0.0419 -	ISM02.4	PPB	6/21/2017 12:55	SMPL
I3772-03	MA4X39	-102	-0.0509 -	ISM02.4	PPB	6/21/2017 12:57	SMPL
I3772-04	MA4X40	-138	-0.0598 -	ISM02.4	PPB	6/21/2017 12:59	SMPL
I3772-05	MA4X41	-130	-0.0578 -	ISM02.4	PPB	6/21/2017 13:02	SMPL
I3772-06	MA4X42	-132	-0.0583 -	ISM02.4	PPB	6/21/2017 13:04	SMPL
I3772-07	MA4X42D	-106	-0.0519 -	ISM02.4	PPB	6/21/2017 13:06	SMPL
I3772-08	MA4X42S	3655	0.8813 -	ISM02.4	PPB	6/21/2017 13:08	SMPL
I3772-09	MA4X50	-110	-0.0529 -	ISM02.4	PPB	6/21/2017 13:11	SMPL
I3772-10	MA4X51	-76	-0.0444 -	ISM02.4	PPB	6/21/2017 13:13	SMPL
I3772-11	MA4X60	-127	-0.0571 -	ISM02.4	PPB	6/21/2017 13:15	SMPL
I3772-12	MA4X61	-99	-0.0501 -	ISM02.4	PPB	6/21/2017 13:18	SMPL
I3772-13	MA4X62	-130	-0.0578 -	ISM02.4	PPB	6/21/2017 13:20	SMPL
I3772-14	MA4X63	-109	-0.0526 -	ISM02.4	PPB	6/21/2017 13:22	SMPL
I3772-15	MA4X64	-60	-0.0404 -	ISM02.4	PPB	6/21/2017 13:24	SMPL
I3772-16	MA4X65	-54	-0.039 -	ISM02.4	PPB	6/21/2017 13:27	SMPL
I3772-17	MA4X66	-38	-0.035 -	ISM02.4	PPB	6/21/2017 13:29	SMPL
CCV011	CCV011	19928	4.9187 -	ISM02.4	PPB	6/21/2017 13:31	SMPL
CCB011	CCB011	20	-0.0206 -	ISM02.4	PPB	6/21/2017 13:34	SMPL
I3772-18	MA4X81	-133	-0.0586 -	ISM02.4	PPB	6/21/2017 13:36	SMPL
I3772-19	MA4X82	-137	-0.0595 -	ISM02.4	PPB	6/21/2017 13:38	SMPL
I3772-20	MA4X83	-135	-0.0591 -	ISM02.4	PPB	6/21/2017 13:40	SMPL
I3772-21	MA4X84	-123	-0.0561 -	ISM02.4	PPB	6/21/2017 13:43	SMPL
CCV012	CCV012	19744	4.873 -	ISM02.4	PPB	6/21/2017 13:45	SMPL
CCB012	CCB012	44	-0.0146 -	ISM02.4	PPB	6/21/2017 13:47	SMPL