



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88251

Temp.: 21.2 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/21/2017	14:15
2	CAL2	1	Water			7.00	06/21/2017	14:16
3	CAL3	1	Water			10.03	06/21/2017	14:19
4	ICV	1	Water			7.01	06/21/2017	14:20
5	CCV1	1	Water			2.01	06/21/2017	14:21
6	I3808-05	1	Water			8.12	06/21/2017	14:22
7	I3808-05DUP	1	Water			8.13	06/21/2017	14:24
8	CCV2	1	Water			12.01	06/21/2017	14:30

LB 88251

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITYw i3808 WorkList ID : 99887 Date : 6/21/2017 1:55:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/21/2017	Water	I3808-05 -1	Corrosivity	Cool 4 deg C	PSEG01	N43	DRUM-001-003-004-COMP	06/21/2017	9040C

Date/Time 06.21.17 1:55 PM
Received by: JP
Relinquished by: HL

Date/Time 06.21.17 4:00 PM
Received by: HL
Relinquished by: JP