



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By :

Run Number: LB88252

Temp.: 21.2 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/21/2017	14:15
2	CAL2	1	Water			7.00	06/21/2017	14:16
3	CAL3	1	Water			10.03	06/21/2017	14:19
4	ICV	1	Water			7.01	06/21/2017	14:20
5	CCV1	1	Water			2.01	06/21/2017	14:21
6	I3808-04	1	Water			8.12	06/21/2017	14:22
7	I3808-04DUP	1	Water			8.13	06/21/2017	14:24
8	I3833-01	1	Water			4.72	06/21/2017	14:26
9	CCV2	1	Water			12.01	06/21/2017	14:30

LB 88252

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i3808

WorkList ID : 99888

Date : 6/21/2017 1:56:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/21/2017	Water	I3808-04 -D	pH	Cool 4 deg C	PSEG01	N43	DRUM-001-003-004-COMP	06/21/2017	9045C
	Water	I3833-01 -D	pH	Cool 4 deg C	STER14	N21	POOLWATER	06/19/2017	SM 4500-PH B

Date/Time 06-21-17 1:55 PM
Received by: JP
Relinquished by: AL

Date/Time 06-21-17 4:30 PM
Received by: AL
Relinquished by: JP