



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88259

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/21/2017	16:00
2	CAL2	1	Water			6.99	06/21/2017	16:01
3	CAL3	1	Water			10.03	06/21/2017	16:03
4	ICV	1	Water			7.01	06/21/2017	16:05
5	CCV1	1	Water			2.01	06/21/2017	16:06
6	I3808-09	1	Solid	20.03	20	9.37	06/21/2017	16:07
7	I3840-09	1	Solid	20.05	20	7.80	06/21/2017	16:10
8	I3840-10	1	Solid	20.02	20	8.09	06/21/2017	16:11
9	I3840-11	1	Solid	20.03	20	8.36	06/21/2017	16:12
10	I3840-12	1	Solid	20.04	20	7.75	06/21/2017	16:14
11	I3840-12DUP	1	Solid	20.03	20	7.76	06/21/2017	16:15
12	CCV2	1	Water			12.02	06/21/2017	16:19

2388239

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3808

WorkList ID : 99886

Date : 6/21/2017 1:54:29 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2017	Solid	I3808-09	Corrosivity		PSEG01		DRUM-002-COMP	06/21/2017	9045C
	Solid	I3840-09	Corrosivity	Cool 4 deg C	ENTA04	N41	WC-B48-B49-002	06/20/2017	9045C
	Solid	I3840-10	Corrosivity	Cool 4 deg C	ENTA04	N41	WC-B53-B54-002	06/20/2017	9045C
	Solid	I3840-11	Corrosivity	Cool 4 deg C	ENTA04	N41	WC-B45-B44-002	06/20/2017	9045C
	Solid	I3840-12	Corrosivity	Cool 4 deg C	ENTA04	N41	WC-B56-B57-002	06/20/2017	9045C

Date/Time

Received by:

Relinquished by:

06/21/17 2:00 PM

J1

CP

Date/Time

Received by:

Relinquished by:

06/21/17 6:00 PM

CP

JP