



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88260

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/21/2017	16:00
2	CAL2	1	Water			6.99	06/21/2017	16:01
3	CAL3	1	Water			10.03	06/21/2017	16:03
4	ICV	1	Water			7.01	06/21/2017	16:05
5	CCV1	1	Water			2.02	06/21/2017	16:50
6	I3844-01	1	Solid	20.03	20	6.08	06/21/2017	16:51
7	I3844-01DUP	1	Solid	20.04	20	6.09	06/21/2017	16:52
8	CCV2	1	Water			12.01	06/21/2017	17:00

1388260

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3840

WorkList ID : 99892

Date : 6/21/2017 4:33:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3844-01	pH	Cool 4 deg C	TULL02	N42	FK-PS-01-008	06/21/2017	9045C

Date/Time 06/21/17 11:33 AM
Received by: JV
Relinquished by: CP

Date/Time 06/21/17 7:00 AM
Received by: CP
Relinquished by: JV