


7471A

LB88262
INSTRUMENT ID: CV4

Linear

 μ Abs.:

42040



A= 0.0000e+000

B= 2.3805e-004

C= -5.7085e-002

Rho= 0.9999260

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	180	0.000	180				
0.2	0.200	0.190	-0.010	1039	0.0 %	1039				
2.5	2.500	2.534	0.034	10883	0.0 %	10883				
5.0	5.000	4.959	-0.041	21073	0.0 %	21073				
7.5	7.500	7.581	0.081	32086	0.0 %	32086				
10.0	10.000	9.950	-0.050	42040	0.0 %	42040				

LB88262 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 21 Jun 2017 17:49:01				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	180	-	0.0000	7471A	PPB	21 Jun 2017 17:50:33	S Std
0.2 - 1	50.2	1039	-	0.2000	7471A	PPB	21 Jun 2017 17:52:50	S Std
2.5 - 1	52.5	10883	-	2.5000	7471A	PPB	21 Jun 2017 17:55:08	S Std
5.0 - 1	55	21073	-	5.0000	7471A	PPB	21 Jun 2017 17:57:25	S Std
7.5 - 1	57.5	32086	-	7.5000	7471A	PPB	21 Jun 2017 17:59:42	S Std
10.0 - 1	510	42040	-	10.0000	7471A	PPB	21 Jun 2017 18:04:32	S Std
ICV10 - 1	ICV10	16024	3.7574	-	7471A	PPB	21 Jun 2017 18:07:08	U SMPL
ICB10 - 1	ICB10	123	-0.0278	-	7471A	PPB	21 Jun 2017 18:09:23	U SMPL
CCV13 - 1	CCV13	21858	5.1461	-	7471A	PPB	21 Jun 2017 18:11:39	U SMPL
CCB13 - 1	CCB13	128	-0.0266	-	7471A	PPB	21 Jun 2017 18:13:54	U SMPL
CRA - 1	CRA	1012	0.1838	-	7471A	PPB	21 Jun 2017 18:16:10	U SMPL
HighStd - 1	HighStd	42307	10.0139	-	7471A	PPB	21 Jun 2017 18:18:26	U SMPL
ChkStd - 1	ChkStd	28730	6.7820	-	7471A	PPB	21 Jun 2017 18:20:42	U SMPL
PB99976BL - 1	PBS	148	-0.0219	-	7471A	PPB	21 Jun 2017 18:22:58	U SMPL
PB99976BS - 1	LCSS	14586	3.4150	-	7471A	PPB	21 Jun 2017 18:25:15	U SMPL
I3781-01 - 1	SB-09-30.5-31.0	344	0.0248	-	7471A	PPB	21 Jun 2017 18:27:32	U SMPL
I3781-02 - 1	SB-09-30.5-31.0-DUP	277	0.0089	-	7471A	PPB	21 Jun 2017 18:29:49	U SMPL
I3781-04 - 1	SB-01-14.5-15.0	268	0.0067	-	7471A	PPB	21 Jun 2017 18:32:07	U SMPL
I3784-01 - 1	P.56-A-COMP	13249	3.0968	-	7471A	PPB	21 Jun 2017 18:34:25	U SMPL
I3784-03 - 1	P.56-B-COMP	11352	2.6452	-	7471A	PPB	21 Jun 2017 18:36:41	U SMPL
CCV14 - 1	CCV14	21492	5.0590	-	7471A	PPB	21 Jun 2017 18:38:56	U SMPL
CCB14 - 1	CCB14	142	-0.0233	-	7471A	PPB	21 Jun 2017 18:41:13	U SMPL
I3784-05 - 1	P.56-C-COMP	449	0.0498	-	7471A	PPB	21 Jun 2017 18:43:28	U SMPL
I3792-01 - 1	OR-03-061917	7062	1.6240	-	7471A	PPB	21 Jun 2017 18:45:44	U SMPL
I3792-01DUP - 1	OR-03-061917DUP	6520	1.4950	-	7471A	PPB	21 Jun 2017 18:48:00	U SMPL
I3792-01MS - 1	OR-03-061917MS	21979	5.1749	-	7471A	PPB	21 Jun 2017 18:50:16	U SMPL
I3792-01MSD - 1	OR-03-061917MSD	20527	4.8293	-	7471A	PPB	21 Jun 2017 18:52:33	U SMPL
I3799-01 - 1	SB-02-7.5-8.0	263	0.0055	-	7471A	PPB	21 Jun 2017 18:54:50	U SMPL
I3799-03 - 1	SB-03-9.5-10.0	406	0.0396	-	7471A	PPB	21 Jun 2017 18:57:07	U SMPL
I3799-05 - 1	SB-04-14.5-15.0	280	0.0096	-	7471A	PPB	21 Jun 2017 18:59:25	U SMPL
I3799-07 - 1	SB-05-9.5-10.0	250	0.0024	-	7471A	PPB	21 Jun 2017 19:01:43	U SMPL
I3799-09 - 1	SB-06-9.5-10.0	267	0.0065	-	7471A	PPB	21 Jun 2017 19:03:59	U SMPL
CCV15 - 1	CCV15	21953	5.1687	-	7471A	PPB	21 Jun 2017 19:06:15	U SMPL
CCB15 - 1	CCB15	190	-0.0119	-	7471A	PPB	21 Jun 2017 19:08:31	U SMPL
I3799-11 - 1	SB-07-12.5-13.0	327	0.0208	-	7471A	PPB	21 Jun 2017 19:10:48	U SMPL
I3799-13 - 1	SB-08-11.5-12.0	323	0.0198	-	7471A	PPB	21 Jun 2017 19:13:04	U SMPL
I3800-01 - 1	CRUSHED-STONE-DUST	1215	0.2321	-	7471A	PPB	21 Jun 2017 19:15:21	U SMPL
I3802-01 - 1	OK-02-062017	2050	0.4309	-	7471A	PPB	21 Jun 2017 19:17:37	U SMPL
I3792-01LX5 - 1		1428	0.2828	-	7471A	PPB	21 Jun 2017 19:19:54	U SMPL
I3792-01A - 1		17683	4.1523	-	7471A	PPB	21 Jun 2017 19:22:10	U SMPL
CCV16 - 1	CCV16	21623	5.0902	-	7471A	PPB	21 Jun 2017 19:24:28	U SMPL
CCB16 - 1	CCB16	170	-0.0166	-	7471A	PPB	21 Jun 2017 19:26:45	U SMPL