


7470A

LB88270
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35126



A= 0.0000e+000

B= 2.8383e-004

C= -2.3342e-002

Rho= 0.9998981

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	80	0.000	80				
0.2	0.200	0.195	-0.005	771	0.0 %	771				
2.5	2.500	2.437	-0.063	8667	0.0 %	8667				
5.0	5.000	5.095	0.095	18033	0.0 %	18033				
7.5	7.500	7.527	0.027	26601	0.0 %	26601				
10.0	10.000	9.947	-0.053	35126	0.0 %	35126				

LB88270 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Jun 2017 10:55:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	80	-	0.0000	7470A	PPB	22 Jun 2017 11:04:52	S	Std
0.2 - 1	50.2	771	-	0.2000	7470A	PPB	22 Jun 2017 11:07:00	S	Std
2.5 - 1	52.5	8667	-	2.5000	7470A	PPB	22 Jun 2017 11:09:08	S	Std
5.0 - 1	55	18033	-	5.0000	7470A	PPB	22 Jun 2017 11:11:16	S	Std
7.5 - 1	57.5	26601	-	7.5000	7470A	PPB	22 Jun 2017 11:13:24	S	Std
10.0 - 1	510	35126	-	10.0000	7470A	PPB	22 Jun 2017 11:22:51	S	Std
ICV11 - 1	ICV11	13384	3.7755	-	7470A	PPB	22 Jun 2017 11:25:21	U	SMPL
ICB11 - 1	ICB11	325	0.0689	-	7470A	PPB	22 Jun 2017 11:27:27	U	SMPL
CCV17 - 1	CCV17	17807	5.0309	-	7470A	PPB	22 Jun 2017 11:29:33	U	SMPL
CCB17 - 1	CCB17	312	0.0652	-	7470A	PPB	22 Jun 2017 11:31:40	U	SMPL
CRA - 1	CRA	751	0.1898	-	7470A	PPB	22 Jun 2017 11:33:46	U	SMPL
HighStd - 1	HighStd	35278	9.9897	-	7470A	PPB	22 Jun 2017 11:35:53	U	SMPL
ChkStd - 1	ChkStd	24552	6.9453	-	7470A	PPB	22 Jun 2017 11:38:00	U	SMPL
PB100005BL - 1	PBW	-5	-0.0248	-	7470A	PPB	22 Jun 2017 11:40:08	U	SMPL
PB100005BS - 1	LCSW	12497	3.5237	-	7470A	PPB	22 Jun 2017 11:42:15	U	SMPL
PB100005TB - 1	PB99961TB	0	-0.0233	-	7470A	PPB	22 Jun 2017 11:44:23	U	SMPL
I3792-03 - 1	OR-03-061917	19	-0.0179	-	7470A	PPB	22 Jun 2017 11:46:31	U	SMPL
I3792-03DUP - 1	OR-03-061917DUP	23	-0.0168	-	7470A	PPB	22 Jun 2017 11:48:39	U	SMPL
I3792-03MS - 1	OR-03-061917MS	12664	3.5711	-	7470A	PPB	22 Jun 2017 11:50:48	U	SMPL
I3792-03MSD - 1	OR-03-061917MSD	13128	3.7028	-	7470A	PPB	22 Jun 2017 11:52:55	U	SMPL
CCV18 - 1	CCV18	17670	4.9920	-	7470A	PPB	22 Jun 2017 11:55:02	U	SMPL
CCB18 - 1	CCB18	319	0.0672	-	7470A	PPB	22 Jun 2017 11:57:09	U	SMPL
I3799-16 - 1	SB-02-COMP	120	0.0107	-	7470A	PPB	22 Jun 2017 11:59:16	U	SMPL
I3799-17 - 1	SB-03-COMP	92	0.0028	-	7470A	PPB	22 Jun 2017 12:01:23	U	SMPL
I3799-18 - 1	SB-04-COMP	83	0.0002	-	7470A	PPB	22 Jun 2017 12:03:29	U	SMPL
I3799-19 - 1	SB-05-COMP	83	0.0002	-	7470A	PPB	22 Jun 2017 12:05:36	U	SMPL
I3799-20 - 1	SB-06-COMP	48	-0.0097	-	7470A	PPB	22 Jun 2017 12:07:43	U	SMPL
I3799-21 - 1	SB-07-COMP	255	0.0490	-	7470A	PPB	22 Jun 2017 12:09:51	U	SMPL
I3799-22 - 1	SB-08-COMP	203	0.0343	-	7470A	PPB	22 Jun 2017 12:11:59	U	SMPL
I3802-03 - 1	OK-02-062017	203	0.0343	-	7470A	PPB	22 Jun 2017 12:14:07	U	SMPL
I3792-03LX5 - 1		62	-0.0057	-	7470A	PPB	22 Jun 2017 12:16:15	U	SMPL
I3792-03A - 1		12986	3.6625	-	7470A	PPB	22 Jun 2017 12:18:22	U	SMPL
CCV19 - 1	CCV19	17689	4.9974	-	7470A	PPB	22 Jun 2017 12:20:29	U	SMPL
CCB19 - 1	CCB19	322	0.0681	-	7470A	PPB	22 Jun 2017 12:22:36	U	SMPL