

ISM02.4


LB88280

INSTRUMENT ID: CV4

Linear

 μ Abs.:

43423



A= 0.0000e+000

B= 2.3065e-004

C= -3.0635e-002

Rho= 0.9999664

Accept=Accepted

Slope
 $y = \text{intercept}$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	-0.026	-0.026	18	0.000	18					
0.2	0.200	0.187	-0.013	945	0.0 %	945					-7
2.5	2.500	2.524	0.024	11074	0.0 %	11074					1
5.0	5.000	5.056	0.056	22055	0.0 %	22055					1
7.5	7.500	7.474	-0.026	32537	0.0 %	32537					0
10.0	10.000	9.985	-0.015	43423	0.0 %	43423					0

NB
 6/22/15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	18	-		0 ISM02.4	PPB	6/22/2017 13:36	Std
	0.2 S0.2	945	-		0.2 ISM02.4	PPB	6/22/2017 13:38	Std
	2.5 S2.5	11074	-		2.5 ISM02.4	PPB	6/22/2017 13:41	Std
	5 S5	22055	-		5 ISM02.4	PPB	6/22/2017 13:43	Std
	7.5 S7.5	32537	-		7.5 ISM02.4	PPB	6/22/2017 13:45	Std
	10 S10	43423	-		10 ISM02.4	PPB	6/22/2017 13:48	Std
ICV	ICV	17446	3.9933 -		ISM02.4	PPB	6/22/2017 13:50	SMPL
ICB	ICB	-21	-0.0355 -		ISM02.4	PPB	6/22/2017 13:53	SMPL
CCV020	CCV020	22373	5.1298 -		ISM02.4	PPB	6/22/2017 13:55	SMPL
CCB020	CCB020	-29	-0.0373 -		ISM02.4	PPB	6/22/2017 13:57	SMPL
PB100006BL	PBS004	-23	-0.0359 -		ISM02.4	PPB	6/22/2017 13:59	SMPL
I3677-01	MDAE00	1390	0.29 -		ISM02.4	PPB	6/22/2017 14:02	SMPL
I3677-02	MDAE00D	1151	0.2348 -		ISM02.4	PPB	6/22/2017 14:04	SMPL
I3677-03	MDAE00S	12656	2.8885 -		ISM02.4	PPB	6/22/2017 14:09	SMPL
I3677-04	MDAE01	1759	0.3751 -		ISM02.4	PPB	6/22/2017 14:11	SMPL
I3677-05	MDAE02	859	0.1675 -		ISM02.4	PPB	6/22/2017 14:13	SMPL
I3677-06	MDAE03	987	0.197 -		ISM02.4	PPB	6/22/2017 14:16	SMPL
I3677-07	MDAE51	1716	0.3652 -		ISM02.4	PPB	6/22/2017 14:18	SMPL
I3677-09	MDADZ0	9647	2.1945 -		ISM02.4	PPB	6/22/2017 14:23	SMPL
PB100007BL	PBS004	-29	-0.0373 -		ISM02.4	PPB	6/22/2017 14:25	SMPL
I3720-01	MDAE32	11337	2.5843 -		ISM02.4	PPB	6/22/2017 14:27	SMPL
I3720-02	MDAE33	399618	HIGH -		ISM02.4	PPB	6/22/2017 14:29	
I3720-03	MDAE34	1489	0.3128 -		ISM02.4	PPB	6/22/2017 14:31	SMPL
I3720-04	MDAE35	662	0.1221 -		ISM02.4	PPB	6/22/2017 14:34	SMPL
I3720-05	MDAE36	1124	0.2286 -		ISM02.4	PPB	6/22/2017 14:36	SMPL
I3720-06	MDAE37	1492	0.3135 -		ISM02.4	PPB	6/22/2017 14:38	SMPL
I3720-07	MDAE39	1038	0.2088 -		ISM02.4	PPB	6/22/2017 14:40	SMPL
I3720-08	MDAE40	1443	0.3022 -		ISM02.4	PPB	6/22/2017 14:43	SMPL
I3720-09	MDAE41	400	0.0616 -		ISM02.4	PPB	6/22/2017 14:45	SMPL
CCV021	CCV021	21954	5.0331 -		ISM02.4	PPB	6/22/2017 14:47	SMPL
CCB021	CCB021	-42	-0.0403 -		ISM02.4	PPB	6/22/2017 14:50	SMPL
I3720-10	MDAE42	32	-0.0233 -		ISM02.4	PPB	6/22/2017 14:52	SMPL
I3720-11	MDAE43	265	0.0305 -		ISM02.4	PPB	6/22/2017 14:54	SMPL
I3720-12	MDAE43D	237	0.024 -		ISM02.4	PPB	6/22/2017 14:56	SMPL
I3720-13	MDAE43S	11998	2.7367 -		ISM02.4	PPB	6/22/2017 14:59	SMPL
I3720-14	MDAE44	417	0.0655 -		ISM02.4	PPB	6/22/2017 15:01	SMPL
I3720-15	MDAE45	348	0.0496 -		ISM02.4	PPB	6/22/2017 15:03	SMPL
I3720-16	MDAE46	584	0.1041 -		ISM02.4	PPB	6/22/2017 15:05	SMPL
PB100008BL	PBS004	-9	-0.0327 -		ISM02.4	PPB	6/22/2017 15:08	SMPL
I3755-02	MC0AD3	1791	0.3825 -		ISM02.4	PPB	6/22/2017 15:10	SMPL
I3755-03	MC0AD4	497	0.084 -		ISM02.4	PPB	6/22/2017 15:12	SMPL
I3755-04	MC0AD5	511	0.0872 -		ISM02.4	PPB	6/22/2017 15:15	SMPL
I3755-05	MC0AD6	388	0.0589 -		ISM02.4	PPB	6/22/2017 15:17	SMPL
I3755-06	MC0AD7	336	0.0469 -		ISM02.4	PPB	6/22/2017 15:19	SMPL
I3755-07	MC0AD8	598	0.1073 -		ISM02.4	PPB	6/22/2017 15:21	SMPL
I3755-08	MC0AD9	1009	0.2021 -		ISM02.4	PPB	6/22/2017 15:24	SMPL

LB88280
INSTRUMENT ID : CV1

I3755-09	MCOAE0	938	0.1857 -	ISM02.4	PPB	6/22/2017 15:26 SMPL
I3755-10	MCOAE1	304	0.0395 -	ISM02.4	PPB	6/22/2017 15:28 SMPL
I3755-11	MCOAE2	2369	0.5158 -	ISM02.4	PPB	6/22/2017 15:30 SMPL
I3755-12	MCOAE3	623	0.1131 -	ISM02.4	PPB	6/22/2017 15:33 SMPL
I3755-13	MCOAE3D	761	0.1449 -	ISM02.4	PPB	6/22/2017 15:35 SMPL
CCV022	CCV022	21512	4.9312 -	ISM02.4	PPB	6/22/2017 15:37 SMPL
CCB022	CCB022	-21	-0.0355 -	ISM02.4	PPB	6/22/2017 15:40 SMPL
I3755-14	MCOAE3S	12215	2.7868 -	ISM02.4	PPB	6/22/2017 15:42 SMPL
I3677-08	MDADY9	10116	2.3026 -	ISM02.4	PPB	6/22/2017 15:46 SMPL
I3720-02DLX10	MDAE33	56313	12.9581 -	ISM02.4	PPB	6/22/2017 16:04 SMPL
I3720-02DL2X50	MDAE33	11548	2.6329 -	ISM02.4	PPB	6/22/2017 16:11 SMPL
CCV023	CCV023	21451	4.9171 -	ISM02.4	PPB	6/22/2017 16:18 SMPL
CCB023	CCB023	-23	-0.0359 -	ISM02.4	PPB	6/22/2017 16:20 SMPL