

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB88285

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP55256
calibration std. phosphate 1 ppm	WP55875
calibration std. phosphate 0.5 ppm	WP55874
calibration std. phosphate 0.3 ppm	WP55873
calibration std. 0 ppm	WP55870
phosphate ICV std.	WP55877
phosphate CCV std.	WP55876
calibration std. phosphate 0.1 ppm	WP55872
calibration std. phosphate 0.05 ppm	WP55871
Combined reagent	WP55879
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
Ascorbic Acid	WP55878

Intercept: -0.0014Slope: 0.6515Regression: 0.999851

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	06/22/2017	15:30
2	CAL2	0.05	1	50	50	0.032	06/22/2017	15:30
3	CAL3	0.10	1	50	50	0.065	06/22/2017	15:31
4	CAL4	0.30	1	50	50	0.186	06/22/2017	15:31
5	CAL5	0.50	1	50	50	0.329	06/22/2017	15:32
6	CAL6	1.00	1	50	50	0.650	06/22/2017	15:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB88285

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.322	0.496	06/22/2017	15:33
2	ICB		1	50	50	0.000	0.002	06/22/2017	15:33
3	CCV1	0.50	1	50	50	0.330	0.509	06/22/2017	15:34
4	CCB1		1	50	50	0.000	0.002	06/22/2017	15:34
5	PB100054BL		1	50	50	0.000	0.002	06/22/2017	15:35
6	PB100054BS	0.50	1	50	50	0.319	0.492	06/22/2017	15:35
7	I3858-01		1	50	50	2.190	3.364	06/22/2017	15:36
8	I3858-01DUP		1	50	50	2.188	3.361	06/22/2017	15:36
9	I3858-01MS	0.50	1	50	50	2.199	3.377	06/22/2017	15:37
10	I3858-01MSD	0.50	1	50	50	2.205	3.387	06/22/2017	15:37
11	I3858-01		200	50	50	0.109	0.169	06/22/2017	16:02
12	I3858-01DUP		200	50	50	0.107	0.166	06/22/2017	16:02
13	I3858-01MS	0.50	200	50	50	0.119	0.185	06/22/2017	16:03
14	I3858-01MSD	0.50	200	50	50	0.120	0.186	06/22/2017	16:03
15	CCV2	0.50	1	50	50	0.320	0.493	06/22/2017	16:04
16	CCB2		1	50	50	0.000	0.002	06/22/2017	16:04