



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88291

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/22/2017	13:00
2	CAL2	1	Water			7.01	06/22/2017	13:01
3	CAL3	1	Water			10.04	06/22/2017	13:03
4	ICV	1	Water			7.00	06/22/2017	13:05
5	CCV1	1	Water			2.01	06/22/2017	13:06
6	I3851-13	1	Solid	20.03	20	9.23	06/22/2017	13:07
7	I3851-13DUP	1	Solid	20.04	20	9.24	06/22/2017	13:10
8	I3851-14	1	Solid	20.01	20	9.00	06/22/2017	13:11
9	I3851-15	1	Solid	20.03	20	10.67	06/22/2017	13:12
10	I3851-16	1	Solid	20.05	20	8.96	06/22/2017	13:13
11	I3851-17	1	Solid	20.06	20	9.74	06/22/2017	13:15
12	I3851-18	1	Solid	20.02	20	7.85	06/22/2017	13:16
13	I3851-19	1	Solid	20.03	20	10.97	06/22/2017	13:18
14	I3851-20	1	Solid	20.04	20	8.42	06/22/2017	13:20
15	I3851-21	1	Solid	20.03	20	10.07	06/22/2017	13:21
16	CCV2	1	Water			12.01	06/22/2017	13:22
17	I3851-22	1	Solid	20.05	20	9.80	06/22/2017	13:24
18	I3851-23	1	Solid	20.01	20	10.63	06/22/2017	13:25
19	I3851-24	1	Solid	20.03	20	10.46	06/22/2017	13:26
20	CCV3	1	Water			12.02	06/22/2017	13:30

10388291

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3851

WorkList ID : 99933

Date : 6/22/2017 12:38:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3851-13	Corrosivity	Cool 4 deg C	TECT01	N42	G-26-0-1.5	06/21/2017	9045C
	Solid	I3851-14	Corrosivity	Cool 4 deg C	TECT01	N42	G-26-1.5-3	06/21/2017	9045C
	Solid	I3851-15	Corrosivity	Cool 4 deg C	TECT01	N42	G-30-0-1.5	06/21/2017	9045C
	Solid	I3851-16	Corrosivity	Cool 4 deg C	TECT01	N42	G-30-1.5-3	06/21/2017	9045C
	Solid	I3851-17	Corrosivity	Cool 4 deg C	TECT01	N42	G-36-0-1.5	06/21/2017	9045C
	Solid	I3851-18	Corrosivity	Cool 4 deg C	TECT01	N42	G-36-1.5-3	06/21/2017	9045C
	Solid	I3851-19	Corrosivity	Cool 4 deg C	TECT01	N42	G-34-0-1.5	06/21/2017	9045C
	Solid	I3851-20	Corrosivity	Cool 4 deg C	TECT01	N42	G-34-1.5-3	06/21/2017	9045C
	Solid	I3851-21	Corrosivity	Cool 4 deg C	TECT01	N42	G-37-0-1.5	06/21/2017	9045C
	Solid	I3851-22	Corrosivity	Cool 4 deg C	TECT01	N42	G-37-1.5-3	06/21/2017	9045C
	Solid	I3851-23	Corrosivity	Cool 4 deg C	TECT01	N42	G-35-0-1.5	06/21/2017	9045C
	Solid	I3851-24	Corrosivity	Cool 4 deg C	TECT01	N42	G-35-1.5-3	06/21/2017	9045C

Date/Time 06/22/17 12:40 PM

Received by: JP

Relinquished by: SR

Date/Time 06/22/17 5:00 PM

Received by: SR

Relinquished by: JP