



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88292

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/22/2017	14:00
2	CAL2	1	Water			7.01	06/22/2017	14:01
3	CAL3	1	Water			10.04	06/22/2017	14:02
4	ICV	1	Water			7.00	06/22/2017	14:03
5	CCV1	1	Water			2.02	06/22/2017	15:05
6	I3852-19	1	Solid	20.03	20	11.18	06/22/2017	15:06
7	I3852-19DUP	1	Solid	20.05	20	11.19	06/22/2017	15:07
8	I3852-20	1	Solid	20.04	20	10.44	06/22/2017	15:09
9	I3852-21	1	Solid	20.01	20	11.11	06/22/2017	15:10
10	I3852-22	1	Solid	20.03	20	8.61	06/22/2017	15:13
11	I3852-23	1	Solid	20.05	20	10.67	06/22/2017	15:14
12	I3852-24	1	Solid	20.04	20	8.78	06/22/2017	15:15
13	I3852-25	1	Solid	20.00	20	9.41	06/22/2017	15:17
14	I3852-26	1	Solid	20.02	20	8.64	06/22/2017	15:19
15	I3852-27	1	Solid	20.04	20	10.49	06/22/2017	15:20
16	CCV2	1	Water			12.02	06/22/2017	15:22
17	I3852-28	1	Solid	20.03	20	8.39	06/22/2017	15:24
18	I3852-29	1	Solid	20.01	20	10.98	06/22/2017	15:25
19	I3852-30	1	Solid	20.03	20	8.75	06/22/2017	15:26
20	I3852-31	1	Solid	20.04	20	10.57	06/22/2017	15:27

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I3852-32	1	Solid	20.04	20	11.15	06/22/2017	15:30
22	I3852-33	1	Solid	20.02	20	11.07	06/22/2017	15:31
23	I3852-34	1	Solid	20.05	20	8.23	06/22/2017	15:32
24	I3852-35	1	Solid	20.03	20	10.67	06/22/2017	15:33
25	I3852-36	1	Solid	20.02	20	10.70	06/22/2017	15:35
26	I3852-36DUP	1	Solid	20.03	20	10.71	06/22/2017	15:37
27	CCV3	1	Water			12.02	06/22/2017	15:39

188292

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 13852

WorkList ID : 99941

Date : 6/22/2017 1:48:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	13852-19	Corrosivity	Cool 4 deg C	TECT01	N42	G-33-0-1.5	06/21/2017	9045C
	Solid	13852-20	Corrosivity	Cool 4 deg C	TECT01	N42	G-33-1.5-3	06/21/2017	9045C
	Solid	13852-21	Corrosivity	Cool 4 deg C	TECT01	N42	G-32-0-1.5	06/21/2017	9045C
	Solid	13852-22	Corrosivity	Cool 4 deg C	TECT01	N42	G-32-1.5-3	06/21/2017	9045C
	Solid	13852-23	Corrosivity	Cool 4 deg C	TECT01	N42	G-31-0-1.5	06/21/2017	9045C
	Solid	13852-24	Corrosivity	Cool 4 deg C	TECT01	N42	G-31-1.5-3	06/21/2017	9045C
	Solid	13852-25	Corrosivity	Cool 4 deg C	TECT01	N42	G-28-0-1.5	06/21/2017	9045C
	Solid	13852-26	Corrosivity	Cool 4 deg C	TECT01	N42	G-28-1.5-3	06/21/2017	9045C
	Solid	13852-27	Corrosivity	Cool 4 deg C	TECT01	N42	G-27-0-1.5	06/21/2017	9045C
	Solid	13852-28	Corrosivity	Cool 4 deg C	TECT01	N42	G-27-1.5-3	06/21/2017	9045C
	Solid	13852-29	Corrosivity	Cool 4 deg C	TECT01	N42	G-39-0-1.5	06/21/2017	9045C
	Solid	13852-30	Corrosivity	Cool 4 deg C	TECT01	N42	G-39-1.5-3	06/21/2017	9045C
	Solid	13852-31	Corrosivity	Cool 4 deg C	TECT01	N42	G-48-0-1.5	06/21/2017	9045C
	Solid	13852-32	Corrosivity	Cool 4 deg C	TECT01	N42	G-48-1.5-3	06/21/2017	9045C
	Solid	13852-33	Corrosivity	Cool 4 deg C	TECT01	N42	G-38-0-1.5	06/21/2017	9045C
	Solid	13852-34	Corrosivity	Cool 4 deg C	TECT01	N42	G-38-1.5-3	06/21/2017	9045C
	Solid	13852-35	Corrosivity	Cool 4 deg C	TECT01	N42	G-50-0-1.5	06/21/2017	9045C
	Solid	13852-36	Corrosivity	Cool 4 deg C	TECT01	N42	G-50-1.5-3	06/21/2017	9045C

Date/Time 06/22/17 12:40P
 Received by: JP
 Relinquished by: JP

Date/Time 06/22/17 6:00PM
 Received by: JP
 Relinquished by: JP