



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88293

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/22/2017	16:40
2	CAL2	1	Water			7.01	06/22/2017	16:41
3	CAL3	1	Water			10.03	06/22/2017	16:42
4	ICV	1	Water			7.01	06/22/2017	16:43
5	CCV1	1	Water			2.01	06/22/2017	16:44
6	I3868-03	1	Solid	20.03	20	6.47	06/22/2017	16:45
7	I3868-03DUP	1	Solid	20.04	20	7.30	06/22/2017	16:46
8	I3868-06	1	Solid	20.05	20	7.30	06/22/2017	16:47
9	I3868-09	1	Solid	20.05	20	8.93	06/22/2017	16:48
10	I3868-12	1	Solid	20.03	20	7.86	06/22/2017	16:49
11	I3868-15	1	Solid	20.01	20	7.02	06/22/2017	16:50
12	I3872-03	1	Solid	20.03	20	9.43	06/22/2017	16:51
13	I3872-06	1	Solid	20.04	20	8.80	06/22/2017	16:52
14	I3872-09	1	Solid	20.03	20	10.27	06/22/2017	16:53
15	CCV2	1	Water			12.02	06/22/2017	17:00

2088293

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3868

WorkList ID : 99951

Date : 6/22/2017 4:31:42 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3868-03	Corrosivity	Cool 4 deg C	PSEG05	N61	CL-01-062217-A	06/22/2017	9045C
	Solid	I3868-06	Corrosivity	Cool 4 deg C	PSEG05	N61	CL-01-062217-B	06/22/2017	9045C
	Solid	I3868-09	Corrosivity	Cool 4 deg C	PSEG05	N61	CL-01-062217-C	06/22/2017	9045C
	Solid	I3868-12	Corrosivity	Cool 4 deg C	PSEG05	N61	CL-01-062217-D	06/22/2017	9045C
	Solid	I3868-15	Corrosivity	Cool 4 deg C	PSEG05	N61	CL-01-062217-E	06/22/2017	9045C
	Solid	I3872-03	Corrosivity	Cool 4 deg C	PSEG05	N61	HR-05-062217-A	06/22/2017	9045C
	Solid	I3872-06	Corrosivity	Cool 4 deg C	PSEG05	N61	HR-05-062217-B	06/22/2017	9045C
	Solid	I3872-09	Corrosivity	Cool 4 deg C	PSEG05	N61	HR-05-062217-C	06/22/2017	9045C

Date/Time 06/22/17 4:31 PM
Received by: JP
Relinquished by: SR

Date/Time 06/22/17 7:00 PM
Received by: SR
Relinquished by: JP