



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88294

Temp.: 21.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/22/2017	17:10
2	CAL2	1	Water			6.99	06/22/2017	17:11
3	CAL3	1	Water			10.02	06/22/2017	17:12
4	ICV	1	Water			7.00	06/22/2017	17:13
5	CCV1	1	Water			2.01	06/22/2017	17:14
6	I3867-02	1	Water			7.31	06/22/2017	17:15
7	I3867-02DUP	1	Water			7.32	06/22/2017	17:16
8	CCV2	1	Water			12.02	06/22/2017	17:18

2088296

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 13867 WorkList ID : 99952 Date : 6/22/2017 4:32:37 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	13867-02	Corrosivity	Cool 4 deg C	PSEG01	N61	A6229	06/22/2017	9040C

Date/Time 06/22/17 4:33 PM
Received by: JP
Relinquished by: JP

Date/Time 06/22/17 7:00 PM
Received by: SR
Relinquished by: JP