

7471A


LB 88305

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36110



A= 0.0000e+000

B= 2.7755e-004

C= -5.6275e-002

Rho= 0.9999780

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.024	-0.024	116	0.000	116				
0.2	0.200	0.193	-0.007	898	0.0 %	898				
2.5	2.500	2.527	0.027	9308	0.0 %	9308				
5.0	5.000	5.006	0.006	18239	0.0 %	18239				
7.5	7.500	7.532	0.032	27341	0.0 %	27341				
10.0	10.000	9.966	-0.034	36110	0.0 %	36110				

LB88305 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 23 Jun 2017 08:21:38

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	116	-	0.0007471A	PPB		23 Jun 2017 08:33:34	S	Std
0.2 - 1	50.2	898	-	0.20007471A	PPB		23 Jun 2017 08:35:52	S	Std
2.5 - 1	52.5	9308	-	2.50007471A	PPB		23 Jun 2017 08:38:10	S	Std
5.0 - 1	55	18239	-	5.00007471A	PPB		23 Jun 2017 08:40:27	S	Std
7.5 - 1	57.5	27341	-	7.50007471A	PPB		23 Jun 2017 08:42:44	S	Std
10.0 - 1	57.5	36110	-	10.00007471A	PPB		23 Jun 2017 08:45:01	S	Std
ICV12 - 1	ICV12	15093	4.1327	-7471A	PPB		23 Jun 2017 08:47:44	U	SMPL
ICB12 - 1	ICB12	192	-0.0030	-7471A	PPB		23 Jun 2017 08:51:28	U	SMPL
CCV24 - 1	CCV24	18612	5.1094	-7471A	PPB		23 Jun 2017 08:58:55	U	SMPL
CCB24 - 1	CCB24	91	-0.0310	-7471A	PPB		23 Jun 2017 09:01:11	U	SMPL
CRA - 1	CRA	1046	0.2340	-7471A	PPB		23 Jun 2017 09:03:27	U	SMPL
HighStd - 1	HighStd	37337	10.3065	-7471A	PPB		23 Jun 2017 09:08:16	U	SMPL
ChkStd - 1	ChkStd	26543	7.3107	-7471A	PPB		23 Jun 2017 09:10:31	U	SMPL
PB100057BL - 1	PBS	-126	-0.0912	-7471A	PPB		23 Jun 2017 09:12:48	U	SMPL
PB100057BS - 1	LCSS	15731	4.3098	-7471A	PPB		23 Jun 2017 09:15:06	U	SMPL
I3808-07 - 1	DRUM-002-COMP	713	0.1416	-7471A	PPB		23 Jun 2017 09:17:23	U	SMPL
I3842-01 - 1	NC-TS-001	3289	0.8566	-7471A	PPB		23 Jun 2017 09:19:40	U	SMPL
I3842-02 - 1	NC-TS-002	3584	0.9385	-7471A	PPB		23 Jun 2017 09:21:58	U	SMPL
I3842-02DUP - 1	NC-TS-002DUP	3368	0.8785	-7471A	PPB		23 Jun 2017 09:24:16	U	SMPL
I3842-02MS - 1	NC-TS-002MS	18933	5.1985	-7471A	PPB		23 Jun 2017 09:26:32	U	SMPL
CCV25 - 1	CCV25	18634	5.1155	-7471A	PPB		23 Jun 2017 09:28:48	U	SMPL
CCB25 - 1	CCB25	75	-0.0355	-7471A	PPB		23 Jun 2017 09:31:04	U	SMPL
I3842-02MSD - 1	NC-TS-002MSD	19051	5.2313	-7471A	PPB		23 Jun 2017 09:33:20	U	SMPL
I3843-01 - 1	E2-Q10-ESW	14892	4.0770	-7471A	PPB		23 Jun 2017 09:35:36	U	SMPL
I3843-02 - 1	E2-Q10-BASE	13369	3.6543	-7471A	PPB		23 Jun 2017 09:37:52	U	SMPL
I3851-01 - 1	G-26-0-1.5	767	0.1566	-7471A	PPB		23 Jun 2017 09:40:09	U	SMPL
I3851-02 - 1	G-26-1.5-3	6022	1.6151	-7471A	PPB		23 Jun 2017 09:42:25	U	SMPL
I3851-03 - 1	G-30-0-1.5	1075	0.2421	-7471A	PPB		23 Jun 2017 09:44:42	U	SMPL
I3851-04 - 1	G-30-1.5-3	12570	3.4325	-7471A	PPB		23 Jun 2017 09:46:59	U	SMPL
I3851-05 - 1	G-36-0-1.5	6621	1.7814	-7471A	PPB		23 Jun 2017 09:49:17	U	SMPL
I3851-06 - 1	G-36-1.5-3	3428	0.8952	-7471A	PPB		23 Jun 2017 09:51:34	U	SMPL
I3851-07 - 1	G-34-0-1.5	1427	0.3398	-7471A	PPB		23 Jun 2017 09:53:51	U	SMPL
CCV26 - 1	CCV26	18971	5.2091	-7471A	PPB		23 Jun 2017 09:56:07	U	SMPL
CCB26 - 1	CCB26	72	-0.0363	-7471A	PPB		23 Jun 2017 09:58:23	U	SMPL
I3851-08 - 1	G-34-1.5-3	8279	2.2415	-7471A	PPB		23 Jun 2017 10:00:40	U	SMPL
I3851-09 - 1	G-37-0-1.5	7183	1.9373	-7471A	PPB		23 Jun 2017 10:02:56	U	SMPL
I3851-10 - 1	G-37-1.5-3	9659	2.6246	-7471A	PPB		23 Jun 2017 10:05:12	U	SMPL
I3851-11 - 1	G-35-0-1.5	2811	0.7239	-7471A	PPB		23 Jun 2017 10:07:29	U	SMPL
I3851-12 - 1	G-35-1.5-3	618	0.1152	-7471A	PPB		23 Jun 2017 10:09:46	U	SMPL
I3864-01DLX10 - 1	1705315435-36-57-COMP	-78	-0.0779	-7471A	PPB		23 Jun 2017 10:12:03	U	SMPL
I3864-02DLX10 - 1	1706055458-59-60-COMP	-11	-0.0593	-7471A	PPB		23 Jun 2017 10:14:20	U	SMPL
I3864-03DLX10 - 1	1706055461-62-71-COMP	14	-0.0524	-7471A	PPB		23 Jun 2017 10:16:38	U	SMPL
PB100058BL - 1	PBS	0	-0.0563	-7471A	PPB		23 Jun 2017 10:18:56	U	SMPL
PB100058BS - 1	LCSS	15229	4.1705	-7471A	PPB		23 Jun 2017 10:21:12	U	SMPL
CCV27 - 1	CCV27	19137	5.2551	-7471A	PPB		23 Jun 2017 10:23:28	U	SMPL
CCB27 - 1	CCB27	-50	-0.0702	-7471A	PPB		23 Jun 2017 10:25:45	U	SMPL
I3852-01 - 1	G-33-0-1.5	3240	0.8430	-7471A	PPB		23 Jun 2017 10:33:02	U	SMPL
I3852-01DUP - 1	G-33-0-1.5DUP	3297	0.8588	-7471A	PPB		23 Jun 2017 10:35:18	U	SMPL
I3852-01MS - 1	G-33-0-1.5MS	19435	5.3378	-7471A	PPB		23 Jun 2017 10:37:34	U	SMPL
I3852-01MSD - 1	G-33-0-1.5MSD	17045	4.6745	-7471A	PPB		23 Jun 2017 10:42:28	U	SMPL
I3852-02 - 1	G-33-1.5-3	390	0.0520	-7471A	PPB		23 Jun 2017 10:44:44	U	SMPL
I3852-03 - 1	G-32-0-1.5	2812	0.7242	-7471A	PPB		23 Jun 2017 10:47:01	U	SMPL
I3852-04 - 1	G-32-1.5-3	13231	3.6159	-7471A	PPB		23 Jun 2017 10:49:19	U	SMPL
I3852-05 - 1	G-31-0-1.5	3384	0.8829	-7471A	PPB		23 Jun 2017 10:51:36	U	SMPL
I3852-06 - 1	G-31-1.5-3	3743	0.9826	-7471A	PPB		23 Jun 2017 10:53:54	U	SMPL
I3852-07 - 1	G-28-0-1.5	1434	0.3417	-7471A	PPB		23 Jun 2017 10:56:11	U	SMPL
CCV28 - 1	CCV28	19502	5.3564	-7471A	PPB		23 Jun 2017 10:58:27	U	SMPL
CCB28 - 1	CCB28	-44	-0.0685	-7471A	PPB		23 Jun 2017 11:00:43	U	SMPL
I3852-08 - 1	G-28-1.5-3	3653	0.9576	-7471A	PPB		23 Jun 2017 11:03:00	U	SMPL
I3852-09 - 1	G-27-0-1.5	1802	0.4439	-7471A	PPB		23 Jun 2017 11:05:16	U	SMPL
I3852-10 - 1	G-27-1.5-3	12021	3.2801	-7471A	PPB		23 Jun 2017 11:07:33	U	SMPL
I3852-11 - 1	G-39-0-1.5	3098	0.8036	-7471A	PPB		23 Jun 2017 11:09:49	U	SMPL
I3852-12 - 1	G-39-1.5-3	6833	1.8402	-7471A	PPB		23 Jun 2017 11:12:06	U	SMPL
I3852-13 - 1	G-48-0-1.5	4286	1.1333	-7471A	PPB		23 Jun 2017 11:14:23	U	SMPL
I3852-14 - 1	G-48-1.5-3	6851	1.8452	-7471A	PPB		23 Jun 2017 11:16:40	U	SMPL
I3852-15 - 1	G-38-0-1.5	1426	0.3395	-7471A	PPB		23 Jun 2017 11:18:58	U	SMPL
I3852-16 - 1	G-38-1.5-3	14796	4.0503	-7471A	PPB		23 Jun 2017 11:21:15	U	SMPL
I3852-17 - 1	G-50-0-1.5	4237	1.1197	-7471A	PPB		23 Jun 2017 11:23:31	U	SMPL
CCV29 - 1	CCV29	19474	5.3487	-7471A	PPB		23 Jun 2017 11:25:47	U	SMPL
CCB29 - 1	CCB29	-48	-0.0696	-7471A	PPB		23 Jun 2017 11:28:04	U	SMPL
I3852-18 - 1	G-50-1.5-3	1624	0.3945	-7471A	PPB		23 Jun 2017 11:30:20	U	SMPL
I3863-01DLX10 - 1	1706125491-92-COMP	10	-0.0535	-7471A	PPB		23 Jun 2017 11:32:37	U	SMPL
I3865-01DLX10 - 1	757005-06-COMP	0	-0.0563	-7471A	PPB		23 Jun 2017 11:34:53	U	SMPL
PB100059BL - 1	PBS	7	-0.0543	-7471A	PPB		23 Jun 2017 11:37:10	U	SMPL
PB100059BS - 1	LCSS	15670	4.2929	-7471A	PPB		23 Jun 2017 11:39:27	U	SMPL
I3842-02LX5 - 1		736	0.1480	-7471A	PPB		23 Jun 2017 11:41:43	U	SMPL
I3842-02A - 1		20654	5.6762	-7471A	PPB		23 Jun 2017 11:44:01	U	SMPL
I3827-01DLX10 - 1	1705245406	-18	-0.0613	-7471A	PPB		23 Jun 2017 11:46:18	U	SMPL
I3827-02DLX10 - 1	1706125495	12	-0.0529	-7471A	PPB		23 Jun 2017 11:48:36	U	SMPL
I3827-03DLX10 - 1	1706125496	21	-0.0504	-7471A	PPB		23 Jun 2017 11:50:53	U	SMPL
CCV30 - 1	CCV30	18718	5.1388	-7471A	PPB		23 Jun 2017 11:53:10	U	SMPL
CCB30 - 1	CCB30	-16	-0.0607	-7471A	PPB		23 Jun 2017 11:55:27	U	SMPL
I3827-04DLX10 - 1	1706055451	20	-0.0507	-7471A	PPB		23 Jun 2017 11:57:44	U	SMPL
I3827-05DLX10 - 1	1705255413	23	-0.0499	-7471A	PPB		23 Jun 2017 12:00:01	U	SMPL
I3828-01DLX10 - 1	1705305429-30-COMP	52	-0.0418	-7471A	PPB		23 Jun 2017 12:02:18	U	SMPL
I3828-02DLX10 - 1	1705305431-32-COMP	31	-0.0477	-7471A	PPB		23 Jun 2017 12:04:35	U	SMPL
I3829-02DLX10 - 1	1706075478	1821	0.4491	-7471A	PPB		23 Jun 2017 12:06:52	U	SMPL
I3830-01DLX10 - 1	1706075480	-141	-0.0954	-7471A	PPB		23 Jun 2017 14:31:03	U	SMPL
I3830-02DLX10 - 1	1706055454	-20	-0.0618	-7471A	PPB		23 Jun 2017 14:36:04	U	SMPL
I3830-03DLX10 - 1	1706075468	-137	-0.0943	-7471A	PPB		23 Jun 2017 14:38:22	U	SMPL
I3830-04DLX10 - 1	1706135404	-91	-0.0815	-7471A	PPB		23 Jun 2017 14:40:39	U	SMPL
I3831-01DLX10 - 1	170605472-73-74-75-COMP	-88	-0.0807	-7471A	PPB		23 Jun 2017 14:42:56	U	SMPL
CCV31 - 1	CCV31	19980	5.4891	-7471A	PPB		23 Jun 2017 14:45:13	U	SMPL

LB88305 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 23 Jun 2017 08:21:38

Sample ID	Extended ID	μ Abs	Conc	Std Cond	Method	Units	Date	Type	Type
CCB31 - 1	CCB31	-192	-0.1096	-7471A	PPB		23 Jun 2017 14:47:30	U	SMPL
I3831-02DLX10 - 1	17K2697	-91	-0.0815	-7471A	PPB		23 Jun 2017 14:49:47	U	SMPL
I3831-03DLX10 - 1	17K2661	-47	-0.0693	-7471A	PPB		23 Jun 2017 14:52:04	U	SMPL
I3831-04DLX10 - 1	17K2721	38	-0.0457	-7471A	PPB		23 Jun 2017 14:54:21	U	SMPL
I3831-05DLX10 - 1	17K2672	-67	-0.0749	-7471A	PPB		23 Jun 2017 14:56:38	U	SMPL
PB100061BL - 1	PBS	180	-0.0063	-7471A	PPB		23 Jun 2017 15:13:26	U	SMPL
PB100061BS - 1	LCSS	15790	4.3262	-7471A	PPB		23 Jun 2017 15:18:23	U	SMPL
CCV32 - 1	CCV32	19455	5.3434	-7471A	PPB		23 Jun 2017 15:20:40	U	SMPL
CCB32 - 1	CCB32	-131	-0.0926	-7471A	PPB		23 Jun 2017 15:22:57	U	SMPL
I3866-01DLX10 - 1	1657002	-58	-0.0724	-7471A	PPB		23 Jun 2017 15:25:44	U	SMPL
CCV33 - 1	CCV33	19297	5.2995	-7471A	PPB		23 Jun 2017 15:28:38	U	SMPL
CCB33 - 1	CCB33	-112	-0.0874	-7471A	PPB		23 Jun 2017 15:30:56	U	SMPL