

Chemtech Consulting Group

Analytical Review Report

Date Printed : 6/26/17
 Analyst : AK
 Data File : LB88309.csv

Approved By : 
 Approved Date :
 Worksheet # : 4/28/17

m9056-A - Inorganic Arsenic - 13

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
			PPB				Method	Method				
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Nitrate												
STD1	STD1				W	6/12/17						
Nitrate		PASS	0.000		0.000						mg/L	
STD2	STD2				W	6/12/17						
Nitrate		PASS	0.190		0.190						mg/L	
STD3	STD3				W	6/12/17						
Nitrate		PASS	0.528		0.528						mg/L	
STD4	STD4				W	6/12/17						
Nitrate		PASS	0.942		0.942						mg/L	
STD5	STD5				W	6/12/17						
Nitrate		PASS	2.443		2.44						mg/L	
STD6	STD6				W	6/12/17						
Nitrate		PASS	4.979		4.98						mg/L	
STD7	STD7				W	6/12/17						
Nitrate		PASS	6.294		6.29						mg/L	
ICV1	ICV1				W	6/12/17						
Nitrate		PASS	2.409		2.41	96.0	90	110			mg/L	
ICB1	ICB1				W	6/12/17						
Nitrate		PASS	0.000		0.000			+/-0.0270			mg/L	
CCV1	CCV1				W	6/24/17						
Nitrate		PASS	2.411		2.41	96.0	90	110			mg/L	
CCB1	CCB1				W	6/24/17						
Nitrate		PASS	0.000		0.000			+/-0.0270			mg/L	
LB88309BLS	LB88309BLS				S	6/24/17						
Nitrate		PASS	0.000		0.000			+/-0.0110			mg/Kg	
LB88309BSS	LB88309BSS				S	6/24/17						
Nitrate		PASS	2.412		48.10	96.0	90.00	110.00			mg/Kg	
I3879-01	I16-005-MH			1	S	6/24/17						
Nitrate		PASS	5.817		128						mg/Kg	
I3879-01D	I16-005-MHD			1	S	6/24/17						
Nitrate		PASS	5.821		128.00				0	20	mg/Kg	
I3879-01S	I16-005-MHS			1	S	6/24/17						
Nitrate		PASS	7.913		174.0	84.0	80	120			mg/Kg	
I3879-01SD	I16-005-MHSD			1	S	6/24/17						
Nitrate		PASS	8.027		177.0	89.0	80	120	1.7	15	mg/Kg	
CCV2	CCV2				W	6/24/17						
Nitrate		PASS	2.422		2.42	97.0	90	110			mg/L	
CCB2	CCB2				W	6/24/17						
Nitrate		PASS	0.000		0.000			+/-0.0270			mg/L	