



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB88310
Analysis Date: 06/23/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 06/23/2017
Extraction IN Time: 14:00
Extraction OUT Time: 14:50

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB88310BL	LB88310BL	WATER	2	1000	100	3.1942	3.1942	3.02	3.1942	3.1942	0	0
2	LB88310BS	LB88310BS	WATER	2	1000	100	3.0611	3.0611	3.03	3.0801	3.0801	0.019	19
3	LB88310BSD	LB88310BSD	WATER	2	1000	100	3.0987	3.0987	3.01	3.1179	3.1179	0.0192	19.2
4	I3808-04	DRUM-001-003-004-COMP	WATER	2	1000	100	3.088	3.088	3.04	3.1245	3.1245	0.0365	36.5
5	I3825-02	GRAB	WATER	2	1000	100	3.108	3.108	3.05	3.1243	3.1243	0.0163	16.3
6	I3867-01	A6229	WATER	2	1000	100	3.1225	3.1225	3.02	3.4951	3.4951	0.3726	372.6



QC Batch# LB88310 Test: TPH Analysis Date: 06/23/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2243
pH Paper 0-14	W2186
Sodium Sulfate	EP1754
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP52823
LCSWD	10.00 ML	WP54439
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 70 Dessicator Time In : 16:00
1.0000 gram Balance: 1.0003 In Time: 15:25
Bal Check Time: 14:25

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 17:30
1.0000 gram Balance: 1.0004 Out Time: 17:00
Bal Check Time: 18:30

1B88310

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I3808

WorkList ID : 100010

Date : 6/23/2017 1:42:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2017	Water	I3808-04-12	TPH	Conc H2SO4 tr	PSEG01	N43	DRUM-001-003-004-COMP	06/21/2017	1664A
	Water	I3825-02	TPH	Conc H2SO4 tr	ARAM01	N23	GRAB	06/21/2017	1664A
	Water	I3867-01-12	TPH	Conc H2SO4 tr	PSEG01	N61	A6229	06/22/2017	1664A

Date/Time 06/23/17 1:42 PM
Received by: 38
Relinquished by: 60

Date/Time 06/23/17 3:30 PM
Received by: 20
Relinquished by: 30